

**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : etenders_ep@kmscl.kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

Equipment Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/99/2018

Dated: 14.06.2018

NOTICE

Sub:- Tender No. T267(R)/2018- Results of technical document evaluation/ demonstration & Date of Price Bid Opening – various equipments- reg.

- Ref:-**
1. Tender Inviting Notice published dated 14.02.2018
 2. Technical bid opening on 12.03.2018 @ 11.00 A M
 3. Demonstration/ document evaluation was conducted on 11.06.2018

I. The reports of technical document evaluation/ demonstration of Various Equipments as per reference no. 3rd cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Model / OEM
1.	Percutaneous Surgical Cricothyroidectomy Set	M/s. Biotrace Impex	C-TCCSB-500,cook medical
2.	PATIENT TRANSFER SLIDE	M/s. universal Agencies	SAFESLIDE
3.	Alpha Bed	M/s. Trivandrum Surgicals and Lab Supplies	AIR CURE 1100
		M/s. Kerala Surgical Equipment Co.	Bioplus
4.	Patient Warmer	M/s. Arkon Medical Systems	Warmtouch 6000
		M/s. JAKESMED HEALTHCARE	Bair hugger 675

II. The items offered by the other tenderer participated in the tender are rejected due to the reasons mentioned against their name below.

A. Patient Warmer : KMSCL/EP/T267(R)/165/2018

1.**M/s.** Kerala Surgical Equipment Co.

Rejected Model/ OEM: KAM3046

Reason for rejection :

Clause No	Technical Specification	Deviation Noted
7	Elapsed Time Display	Not Available
8	IV pole/bedrail mountable	Not Available
	Comment	Very Slow warming and set temperature not attained

Iv. The online price bid of the eligible tenderer mentioned above will be opened on 16.06.2018 @ 11.00 am at the etender portal.

Dr. Dileep Kumar S R
General Manager