



Running Contract Details	
Equipment Name	ACT Machine (Activated Clotting Time)
Running Contract Valid Till	22-06-2020
Tender Ref No	KMSCL/EP/T266/858/2017
Tendered Quantity	25
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	4 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	ACT Machine (Activated Clotting Time) <i>Model & Make : SONOCLOT / LIFE DIAGNOSTICA</i>	223020 Incl.GST :18%	15611.4	238631.4
2	Cost per test	120.96 Incl.GST :12%	8.92	129.88
		223140.96	15620.32	238761.28

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
ACT Machine (Activated Clotting Time)				
Labour	3,540.00	3,540.00		3,540.00
Comprehensi ve	5,310.00	5,900.00		7,080.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :ACT Machine (Activated Clotting Time)

1. Real time clot detection
2. Liquid suspended kaolin activator for uniform mixing with blood sample
3. Room temperature storage
4. voltage: 220-240 V, 50 Hz Single ~ Phase
5. Temperature : 14o C to 32o C (57o F to 90oF)
6. Humidity : 10% to 90%, non condensing
7. Installation and training
8. Must be compact, small & portable
9. The cost per test shall be offered in BOQ and 500 tests will be taken for evaluation. The rate offered shall be fixed for a period of three years.
10. Consumables required for 10 tests shall be provided with the machine
11. Should have display for showing the results.