



Running Contract Details	
Equipment Name	Ablative Carbon Dioxide Laser
Running Contract Valid Till	24-04-2021
Tender Ref No	KMSCL/EP/T315/1263/2019(R)
Tendered Quantity	20
Supplier Name	M/s C Cube Advanced Technologies
GST No	29ANCP4091F1ZS
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
406 3rd floor Malnad Residency 4th Main Hanuma Hills layout Banashankari 3rd stage Bangalore 560061	Contact Person	Vinayaka KS
	Phone	8043021028
	Mobile No	9611111273
	Email	ccubeadvtech@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Ablative Carbon Dioxide Laser <i>Model & Make : carby / C CUBE ADVANCED TECHNOLOGIES</i>	1372000 Incl.GST :12%	101185	1473185
		1372000	101185	1473185

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Ablative Carbon Dioxide Laser							
Labour	65,000.00	65,000.00	65,000.00	65,000.00	65,000.00	65,000.00	65,000.00
Comprehensive	1,15,000.00	1,15,000.00	1,15,000.00	1,15,000.00	1,15,000.00	1,15,000.00	1,15,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Ablative Carbon Dioxide Laser

1. Wavelength : 10600nm
2. Power : between 25-50 watt
3. Modes : continuous, Repeat , super pulse/ultra pulse
4. Spot size: 0.2-0.3 mm
5. Pulses: step wise increase in energy should be possible.
6. Display: LCD /LED
7. Cooling System : needed – water /Air
8. Aiming beam : Between 2-5 mw power
9. Electrical: 220V/50Hz
10. Delivery System : Easily Manourable Articulated Arm
11. Accessories to be supplied : patient eyeshield- 2 sets & Physician goggles-atleast 5 Nos