

Running Contract Details	
Equipment Name	Adult CPR Manikin
Running Contract Valid Till	27-12-2020
Tender Ref No	KMSCL/EP/T294/834B/2018
Tendered Quantity	276
Supplier Name	M/s Laerdal Medical India Pvt Ltd
GST No	33AABCL8105F1ZX
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
No.10 1st Floor 1st Street Dr. Subbarayan Nagar Kodambakam Chennai- 600 024.	Contact Person	Balaganesh
	Phone	044-42614773
	Mobile No	9840153367
	Email	balaganesh.mahendrathas@laederal.co.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Adult CPR Manikin <i>Model & Make : Little Anne QCPR / Laerdal Medical Norway</i>	53100 Incl.GST :18%	3717	56817
		53100	3717	56817

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Adult CPR Manikin

The manikin can be used for training of CPR, with feedback on the performed CPR and with an option to also accommodate training in Rescue/ extraction of life savings first aid in future.

1. The manikin should be designed as per the new 2015 guideline of American Heart Association
2. Airway must open/close by the proper method of head tilt. Chin lift and Jaw thrust
3. Ventilation of the manikin must be possible through the following procedures; mouth to mouth, mouth to nose, Mask to mouth (both packet Mask and Bag- Valve Mask (BVM))
4. The manikin must show a realistic chest rise during ventilation
5. The student must be provided correct feedback for either normal ventilation volumes or for volumes when supplementary oxygen is provided based on the 2015 AHA guidelines
6. Removal and replacement of lungs and face skin must be easy without use of tools
7. The manikin must be equipped with a bilateral carotid pulse that can be manually controlled.
8. The manikin must be equipped with means of providing feedback on performed CPR, compression recoil, depth, rate, rhythm, interpretation time& ventilation volume on a battery operated audio visual electronic feedback unit. Models supplied with feedback unit/manikin that requires 230V power connection will not be accepted
9. The manikin must be available as a Torso model
10. Should provide 100 numbers of face shield along with manikin
11. The individual manikin configuration should come standard with a carry Bag which also serves as a padded training mat.