



Running Contract Details	
Equipment Name	Advanced Electrosurgical unit with bipolar cutting and Coagulation
Running Contract Valid Till	17-01-2021
Tender Ref No	KMSCL/EP/T298/1032B/2018(R)
Tendered Quantity	20
Supplier Name	M/s Erbe Medical India Pvt Ltd
GST No	33AABCE7430L1ZP
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Ground Floor Jayant Tech Park41 Mount Poonamalle High Road Nandambakkam Chennai 600089	Contact Person	Nimin Sukumaran
	Phone	04471555100
	Mobile No	9567771552
	Email	nimin.sukumaran@erbe-india.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Advanced Electrosurgical unit with bipolar cutting and Coagulation</b> <i>Model &amp; Make : VIO S / Erbe Elektromedizin GmbH Submitted</i>	537600 Incl.GST :12%	39648	577248
		<b>537600</b>	<b>39648</b>	<b>577248</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Advanced Electrosurgical unit with bipolar cutting and Coagulation</b>							
Labour	4,800.00	5,280.00	8,640.00	9,504.00	10,464.00	11,520.00	12,480.00
Comprehensive	9,600.00	10,560.00	21,600.00	24,000.00	25,920.00	28,320.00	31,200.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Advanced Electrosurgical unit with bipolar cutting and Coagulation**

1. It should be microprocessor unit with all necessary modalities comprising of Monopolar & Bipolar modes with LCD display and option to store program.
2. It should have Bipolar modes with Auto start and Auto stop.
3. The unit should have facility to upgrade with irrigation pump for future requirement.
4. The unit should also have a special continues neutral electrode monitoring system with equipotential ring plate.
5. System should have safety CE certificate issued by a notified body registered in European commission and FDA (US) certificate
6. The unit should be supplied with all the standards accessories for the hassle free electrosurgical applications.
  - a) Two pedal foot switch
  - b) Nonstick Bayonet forceps -1 No. (0.4 mm/0.7 mm-20 cm/23cm)
  - c) Straight Bipolar Forceps - 2 Nos. (19cm/225 with 1mm punctiform tip)
  - d) Disposable Omega patient plate -1 Box (50 Nos/Box)
  - e) Monopolar Pencil - 1 Box (10 Nos/Box)
  - f) Bipolar cable - 3 Nos