



Running Contract Details	
Equipment Name	Amplitude EEG Machine
Running Contract Valid Till	06-06-2021
Tender Ref No	KMSCL/EP/T324/1391/2019(R)
Tendered Quantity	15
Supplier Name	M/s Axis Health Care
GST No	32AAQFA7508K1Z9
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
TC No. 2/3031/7 Shop No.G6 Dally's Apartments Chalakuzhy Rd Trivandrum -695004	Contact Person	Anil Kumar T V
	Phone	0471-2552806
	Mobile No	9947425200
	Email	axishealthcare@mail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Amplitude EEG Machine <i>Model & Make : Natus Nicolet V 32 / Natus Medical</i>	672000 Incl.GST :12%	49560	721560
2	Additional EEG electrode	156.8 Incl.GST :12%	11.56	168.36
3	Additional Conductivity paste	1512 Incl.GST :12%	111.51	1623.51
4	Additional bottle of Skin Prepping Gel	1344 Incl.GST :12%	99.12	1443.12
5	Amplitude EEG, ICU software package	380800 Incl.GST :12%	28084	408884
6	Digital video EEG software & hardware	425600 Incl.GST :12%	31388	456988
		1481412.8	109254.19	1590666.99

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Amplitude EEG Machine							
Labour	10,000.00	10,000.00	10,000.00	10,000.00	10,000.00	10,000.00	10,000.00
Comprehensive	35,000.00	35,000.00	40,000.00	40,000.00	45,000.00	45,000.00	50,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Amplitude EEG Machine

1. The software should be able to display aEEG (amplitude integrated EEG) which is necessary for ICU Recording
2. Software should have a ability to show power and frequency trending
3. It should have CSA (Compressed Spectral Array)
4. It should have BSI (Brain Symmetry Index) For ICU EEG purpose
5. Should have spike & seizure detection software.
6. Should have digital video EEG Hardware & software
7. Must have minimum 32 referential EEG Channels.
8. Should have standard 10-20 channel labels
9. Sampling rate should be up to 2 KHz.
10. 16 bits of ADC resolution for digitization of signals.
11. Should have Digital EEG software working on windows 10 operating system.
12. Patient Database software should be based in Microsoft SQL server database.
13. Must have onboard impedance check with colored LED's & Impedance button with thresh hold LEDs
14. Should be supplied with a latest Monoputer with a compact trolley for ICU EEG

15. Must be supplied with 100 nos of EEG electrodes, 5 bottle of Conductivity Paste and 5 bottle of Skin Prepping Gel free of cost along with the machine.
16. Must be supplied with digital video EEG software & hardware
17. The system must have CE & FDA
18. Rate for one EEG electrode, one bottle of Conductivity Paste 100 gm and one bottle of Skin Prepping Gel 100 gm should be quoted separately.

For evaluation purpose rate of 1000 electrodes, 10 bottle of Conductivity Paste and 10 bottle of Prepping Gel will be taken and the calculation will be automatically done by the BOQ.

The rate will be fixed for 3 years from the date of price bid opening.