

Running Contract Details	
Equipment Name	Argon Plasma Coagulator
Running Contract Valid Till	22-02-2021
Tender Ref No	KMSCL/EP/T289/1286/2018
Tendered Quantity	25
Supplier Name	M/s Axis Medical Systems
GST No	32AARFA4092K1Z7
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
K.C.No.13/792/3 Kajas Building Annie Hall Road Kozhikode- 673002 Kerala	Contact Person	Jagdeesh.T.S/K.P.Santhosh
	Phone	
	Mobile No	9526600078/9349114142
	Email	axismedical@mail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Argon Plasma Coagulator <i>Model & Make : WEM ARGON 4 PLASMA COAGULTOR / MEDTRONICS PLC</i>	851200 Incl.GST :12%	62776	913976
		851200	62776	913976

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Argon Plasma Coagulator							
Labour	20,088.00	21,093.00	22,147.00	23,255.00	24,417.00	25,638.00	26,920.00
Comprehensive	31,860.00	33,453.00	35,126.00	36,882.00	38,726.00	40,663.00	42,695.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Argon Plasma Coagulator

- I. The unit should have One APC socket to connect APC filter instruments and should be compatible with Electrosurgical Units.
 - II. The APC unit should have four modes of APC – Micro, Macro, Pulsed High and Pulsed Low with possibility to modulate flow and power settings.
 - III. The APC applicator should be filter integrated with option for plug and play and option to change between two modes from a sterile zone.
 - IV. The system should have the option to control the flow of Argon Gas and button to purge the instrument.
 - V. It should come with open and lap probes with spatula electrode, shaft length for open 70 to100mm and Lap probe 100 to 320mm
 - VI. The unit should be supplied with suitable imported cart and accessories for building the workstation.
- i. Argon gas cylinder 5 Ltr. - 2 Nos.
 - ii. Pressure reducer with sensor for APC - 1 Nos.
 - iii. APC probe-Open - 5 Nos.
 - iv. APC probe -Lap - 5 Nos.