

Running Contract Details	
Equipment Name	Autoclave Class B
Running Contract Valid Till	02-04-2021
Tender Ref No	KMSCL/EP/T315/45C/2019(R)
Tendered Quantity	27
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Autoclave Class B <i>Model & Make : TRUEKLAV / Sterile Safequip & Chemicals Ltd</i>	279660 Incl.GST :18%	19576.2	299236.2
		279660	19576.2	299236.2

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Autoclave Class B							
Labour	2,000.00	2,000.00	2,000.00	2,000.00	2,000.00	2,000.00	2,000.00
Comprehensive	4,000.00	4,000.00	4,000.00	4,000.00	4,000.00	4,000.00	4,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Autoclave Class B

1. B type fully automatic loading autoclave, capacity 22 Litre with vacuum, dry cycle & micro-processor controlled wrapped/unwrapped, porous, solid instrument including cotton swabs can be sterilized.
2. Bow & Dick test qualified, inbuilt sensor for distilled water.
3. Digital display of pressure and temperature.
4. Double safety lock on the door, silicon jacket.
5. Diagnostic automatic alarm for any malfunctioning.
6. Can be used at 121o C/ 134o C during sterilization.
7. Two storage tanks for fresh and used water.
8. Chamber size 250 mm x 450 mm.
9. Number of trays :3.
10. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.
11. Should have automatic motorised door locking system
12. Should have inbuilt steam generator
13. Should have automatic cut-off to prevent overheating in chamber