

Running Contract Details	
Equipment Name	Automated Peritoneal Dialysis Cyclor
Running Contract Valid Till	17-06-2021
Tender Ref No	KMSCL/EP/T324/1346/2019(R)
Tendered Quantity	50
Supplier Name	M/s Baxter (India) Pvt Ltd
GST No	
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Plot No: 183 Sector - 5 IMT Manesar Gurgaon Haryana - 122050	Contact Person	Mrs. Zia Mohammad
	Phone	
	Mobile No	9899760197
	Email	zia_mohammad@baxter.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Automated Peritoneal Dialysis Cyclor <i>Model & Make : 5C6M10 / Baxter Healthcare S.A</i>	780595.2 Incl.GST :12%	57568.9	838164.1
		780595.2	57568.9	838164.1

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Automated Peritoneal Dialysis Cyclor							
Labour	19,051.00	20,004.00	21,004.00	22,054.00	23,157.00	24,315.00	25,530.00
Comprehensive	69,696.00	76,665.00	84,332.00	92,765.00	1,02,042.00	1,12,246.00	1,23,471.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Automated Peritoneal Dialysis Cyclers

1. Therapy options available
 - a) CCPD/IPD
 - b) Hi-Dose CCPD
 - c) Tidal
 - d) Hi-Dose Tidal
2. Automatic calculation of number of cycles and Dwell time.
3. Use disposable cassette assembly for therapy.
4. Built in fluid warmer: 35 to 37 Degree C.
5. Maximum total programmable volume: 80L
6. Therapy time: 0 to 48 Hrs.
7. Standard and low fill modes
 - a) Low fill mode fill volume: 60 to 1000ml
 - b) Standard mode fill volume: 200 to 3000ml
8. Self-correcting, continuous and system error alarms for maximized patient safety.
9. Use pneumatic pump to exert pressure for fill/ drain the patient.
10. Built in event log & therapy log.
11. Built in nurses & service Menu.
12. Protective system preventing overflow.
13. Program lock.
14. Last fill with same or different dextrose
15. Mode of operation: continuous.

16. Volumetric accuracy reported: Greater of 1% or +/-10ml.
17. Operating voltage: 220 V.
18. Maximum power consumption: 600VA (Average 100VA)
19. Weight of equipment: 12kg.
20. Air detection capacity: Bubble exceeding 3CC.
21. Battery backup for program: Up to 2 Hours.