

Running Contract Details	
Equipment Name	Automated Stainer for Histopathology Lab
Running Contract Valid Till	21-01-2021
Tender Ref No	KMSCL/EP/T292/1275/2018
Tendered Quantity	15
Supplier Name	M/s Bions Medical Systems Pvt Ltd
GST No	32AACCB4877B1Z1
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
28/3085 D âœRohiniâœ Tagore Nagar Ponneth Temple Road Kadavanthra Kochi-682020.	Contact Person	Jitto John
	Phone	
	Mobile No	9567860501
	Email	bions@vsnl.net

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Automated Stainer for Histopathology Lab</b> <i>Model &amp; Make : ST5010 / Leica Microsystems GmbH</i>	1900000.6 Incl.GST :18%	133000.04	2033000.64
		<b>1900000.6</b>	<b>133000.04</b>	<b>2033000.64</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Automated Stainer for Histopathology Lab</b>							
Labour	48,305.00	50,720.00	53,256.00	55,919.00	58,715.00	61,650.00	64,733.00
Comprehens ive	80,509.00	84,534.00	88,761.00	93,199.00	97,859.00	1,02,752.00	1,07,890.00

### Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Automated Stainer for Histopathology Lab**

1. Programmable fully automatic stainer will be used primarily for routine staining procedures in histology and cytology laboratories.
2. Capable of performing multiple staining procedures randomly or in batch mode.
3. Should have capability of processing upto 1 to 330 slides or more.
4. Should facilitate loading 1 to 30 slides per batch or higher.
5. Should have reagent reservoirs, 450ml each with cover and arranged in two levels.
6. Programmable wash reservoirs of two to five with user defined option.
7. Should have drying station with paraffin waste tray with temperature range 35-65 °C.
8. Should have facility of programming start station and end station.
9. Facility of upto 15 or more programs should be available.
10. Should operate with robotic / similar arm movement and enhanced wash feature.
11. Aspirating fan drawing fumes through filter, optional duct connection adaptor, to be available if required in future.
12. Ability to process in monitor - basket positions.
13. Alarms should be audible beep and screen display messages should be available for user convenience.
14. Diagnostic function like error messages and codes should be available.
15. Safety features like door and cover closure sensors should be available.
16. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.