

Running Contract Details	
Equipment Name	Automatic tissue processor
Running Contract Valid Till	14-10-2021
Tender Ref No	KMSCL/EP/T324/794/2019(R)
Tendered Quantity	15
Supplier Name	M/s Bions Medical Systems Pvt Ltd
GST No	32AACCB4877B1Z1
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
28/3085 D âœRohiniâœ Tagore Nagar Ponneth Temple Road Kadavanthra Kochi-682020.	Contact Person	Jitto John
	Phone	
	Mobile No	9567860501
	Email	bions@vsnl.net

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Automatic tissue processor <i>Model & Make : TP 1020 / Leica Microsystems GmbH</i>	1511555.22 Incl.GST :18%	105808.87	1617364.09
		1511555.22	105808.87	1617364.09

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Automatic tissue processor							
Labour	38,429.00	40,351.00	42,368.00	44,487.00	46,711.00	49,047.00	51,499.00
Comprehens ive	64,049.00	67,251.00	70,614.00	74,145.00	77,852.00	81,744.00	85,832.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Automatic tissue processor

1. Microprocessor controlled Tissue Processor with freely selectable programs
2. User programmable parameters like infiltration time, delay time, agitation on-off and doublee basket operation.
3. Programmable infiltration time.
4. Delayed start up function to a maximum of 9 days.
5. Drain time of 60 second in each station to reduce carry over contamination.
6. Glass Reagent containers of at least 1.8 litres or more with seals to minimize evaporation and exposure the hazardous fumes.
7. Agitation in 3 second intervals with on/off function for through and even mixing of reagents.
8. Maximum tissue processing capacity of 200 or more tissue cassettes using with double basket operation(100/ basket).
9. Maximum safety concept with automatic immersion of the tissue basket into the beaker in case of power failure.
10. Audible alarms, error message and warning codes for maximum safety.
11. Wax bath over temperature (at 650 C) and under temperature cut-off facility for the safety of tissues.
12. Locking facility to avoid inadvertent operation.
13. Facility of manual lifts the carousal and removable tissues in case of long power failures
14. UPS with 60 minutes backup of suitable capacity should be supplied along with the equipment
15. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid. ?