



Running Contract Details	
Equipment Name	AUTOPSY TABLE
Running Contract Valid Till	28-03-2021
Tender Ref No	KMSCL/EP/T308/792/2019
Tendered Quantity	20
Supplier Name	M/s Span Surgical Co
GST No	33AEHPD0152P1ZO
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
305 - Sri Thevar Complex No. 153 - Avanashi Road Coimbatore - 641 018	Contact Person	Vipul Joshi
	Phone	
	Mobile No	9047110541
	Email	spansurgical@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Autopsy Table Model B <i>Model & Make : SP-101 / Span surgical co</i>	462560 Incl.GST :18%	32379.2	494939.2
2	Autopsy Table <i>Model & Make : SP-200 / Span surgical co</i>	678500 Incl.GST :18%	47495	725995
		1141060	79874.2	1220934.2

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Autopsy Table Model B							
Labour	750.00	950.00	1,100.00	1,200.00	1,350.00	1,450.00	1,600.00
Comprehensive	950.00	1,150.00	1,300.00	1,450.00	1,600.00	1,700.00	1,900.00
Autopsy Table							
Labour	750.00	950.00	1,100.00	1,200.00	1,350.00	1,450.00	1,600.00
Comprehensive							

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
ve	950.00	1,150.00	1,300.00	1,450.00	1,600.00	1,700.00	1,900.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Autopsy Table Model B

1. Construction 304 type stainless steel with large radius inside corners for easy clean up
2. Manufactured without rivets, bolts or other devices of that work surface remains smooth and prevent accumulation of bacteria
3. Two piece construction for easy installation and flexibility for complete room wash down
4. Table length 91 inches Width 30 inches Height 33 inches approx.
5. Dissection wing length 62 inches Width 30 inches height 36 inches approx
6. "L" shaped design includes dissection wing with base cabinet
7. Legs provided full access under the table for sanitation and cleaning
8. Three solid stainless steel sliding body support
9. Wooden head rest
10. Base cabinet provided under the table permit additional storage
11. Double bowl sink with rounded corners. Slopping work surface to drain into the integral sink
12. Wrist operated stainless steel faucet with hot and cold water flow control, valves
13. Reverse flow hydro aspirator with built in vacuum breaker
14. Regulated sprinkler system for table surface to ensure continuous self cleaning

Equipment :Autopsy Table

1. Elevating with stainless steel dissection board, Dimension:100"L x 40"W.
2. Elevating height 32" to 40' up and down pedestal control. All thick gauge heavy duty stainless construction with anti corrosion

Elevating/rotating down draft table.

3. Table top and sink fabricate in 16 gauge stainless steel anticorrosion pedestal of 14 gauge stainless steel 304 Grade with satin finish 1/2" removable perforated grid plate with 3/8" diameter hole on 2" centres (4 each). Large Sink well Handy spray, rinse, heavy duty chrome plated hand piece durable 8' long flexible hose Hand piece with hose drop in deck. Concealed pressure control hot and cold water mixture/swing spout Hot/cold mixing valve down draft cones Hydro aspirator with built in vacuum breaker. Access panel 2 outlet curved moulded sides of table (Stainless Steel 304 Grade).

4. Dissection Station 'L' shaped extension. 14 gauge stainless steel 304 Grade extension with drainage facility and water supply (Hot and Cold)

5. (L) 62" X 30" (W) for dissection of internal organs. Extension should have independent fixed pedestal and not attached to autopsy table as pedestal is fixed. The edger should be curved, moulded and raised