



Running Contract Details	
Equipment Name	Baby Warm Pack with bags
Running Contract Valid Till	28-03-2021
Tender Ref No	KMSCL/EP/T311/596/2019(R)
Tendered Quantity	125
Supplier Name	M/s Phoenix Medical Systems Pvt Ltd
GST No	33AAACP1905E1Z2
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
DP 42 SIDCO Industrial Estate Thirumudivakkam Chennai 600 132	Contact Person	Akhil M P
	Phone	
	Mobile No	9388224376
	Email	sales@pmsind.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Baby Warm Pack with bags <i>Model & Make : Embrace Nest Infant Warmer / Phoenix Medical Systems Pvt Ltd</i>	21000 Incl.GST :12%	1548.75	22548.75
		21000	1548.75	22548.75

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Baby Warm Pack with bags				
Labour	850.00	900.00	950.00	
Comprehensive	1,300.00	1,450.00	1,600.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Baby Warm Pack with bags

1. It should be lightweight, portable with Heater and Wrap
2. It should be able to provide warmth to the newborn by the mother's side (in wards) as well as during transport.
3. It should be usable even when there is a power failure
4. It should maintain temperature of approximately 37°C for minimum 4 hours (at an ambient of 25°C).
5. It should not require constant power supply and should be able to provide warmth without electricity.
6. It should have precision heating mechanism to prevent overheating without any action or controls to be set from the user.
7. It should be less than 5 Kg in weight.
8. It should be easily washable with soap and easy to sanitize with disinfectants.
9. Product components that come in contact with patient's skin should be biocompatible.
10. It should provide warmth to patient without having any electricity near the patient, for safety.
11. Power consumption should be less than 550W.
12. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.