

Running Contract Details	
Equipment Name	Bacteriology incubator Model B
Running Contract Valid Till	15-02-2021
Tender Ref No	KMSCL/EP/T299/396B/2018
Tendered Quantity	50
Supplier Name	M/s Labline Instruments
GST No	32AABFL5818N1Z6
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
P B No. 36 Eroor Road Trippunithura-682301	Contact Person	Bini
	Phone	04842776582
	Mobile No	9020304422
	Email	lablinecochin@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Bacteriology incubator Model B <i>Model & Make : BOD-I / LABLINE INSTRUMENTS, COCHIN</i>	67201 Incl.GST :18%	4704.07	71905.07
		67201	4704.07	71905.07

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Bacteriology incubator Model B				
Labour	1,200.00	1,200.00	1,200.00	
Comprehensive	5,000.00	5,000.00	5,000.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Bacteriology incubator Model B

1. Double walled body with inner chamber of stainless steel/CRCA and outer galvanized steel with noncorrosive epoxy powder coated, full view glass door.
2. Capacity: minimum 100 L with internal fan for uniform air circulation
3. Inner chamber: 2-3 shelves and illumination control. Should be available with a switch
4. Temperature: microprocessor based with digital display, range: 5 °C to 50° C with an accuracy of +/- 0.5°C. Uniformity +/-1.0°C throughout the chamber.
5. Door alarm: low/high temperature alarm.
6. Cooling: CFC free refrigeration.
7. Power supply: 210-240V/50-60Hz.
8. Should be double door system, inner door with latch moundable transparent
9. Suitable stabilizer should be provided.
10. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid