

| Running Contract Details | |
|--------------------------------|----------------------------------|
| Equipment Name | Bariatric Operation Table |
| Running Contract Valid Till | 07-08-2020 |
| Tender Ref No | KMSCL/EP/T267/1034/2018(R) |
| Tendered Quantity | 4 |
| Supplier Name | M/s Kerala Surgical Equipment Co |
| GST No | 32AACFK7295D1ZG |
| Installation & Delivery Period | 6 Week(s) |
| Up-time / PM vist | 95% & 4 Visits per year |
| Warranty period | 3 Years |

| Supplier`s Details | | |
|---|-----------------|--------------------------|
| Address | Contact Details | |
| No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18 | Contact Person | Mr. Abbas |
| | Phone | |
| | Mobile No | 9539012460 |
| | Email | info@keralasurgicals.com |

| Item-wise Price Details | | | | |
|-------------------------|--|---|------------------------------------|----------------|
| # | Item Details | Unit Rate (Incl.all taxes & charges) | Service Charges (Through KMSCL) | Grand Total |
| 1 | Bariatric Operation Table <i>Model & Make : PSI / PALAKKAD SURGICALS</i> | 1215400 Incl.GST :18% | 85078 | 1300478 |
| | | 1215400 | 85078 | 1300478 |

| Annual / Comprehensive Maintenance Charges (Exl.Tax) | | | | | | | |
|--|----------------------|----------------------|----------------------|----------------------|----------------------|----------------------|-----------------------|
| Rate | 4 th Year | 5 th Year | 6 th Year | 7 th Year | 8 th Year | 9 th Year | 10 th Year |
| Bariatric Operation Table | | | | | | | |
| Labour | 1,180.00 | 1,180.00 | 1,180.00 | 1,180.00 | 1,180.00 | 1,180.00 | 1,180.00 |
| Comprehensive | 5,900.00 | 5,900.00 | 5,900.00 | 5,900.00 | 5,900.00 | 5,900.00 | 5,900.00 |

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Bariatric Operation Table

1. The five section table top should be 100% X-ray translucent for fluoroscopy with C-Arm with radiolucent mattress and facility to load X-ray cassettes to the table. There should be no metal frame across especially between back and seat section.
2. With remote control the following positions should be possible
 - a. Table up and down
 - b. Table down and up
 - c. Side tilt/lateral left and right
 - d. Trendelenburg and reverse trendelenburg
 - e. Back section up and down
 - f. Flex/reflex positioning should be pre-programmed and therefore should be achievable by press of a single button.
 - g. Powered floors lock and unlock
 - h. Auto levelling
3. Radiolucent five section table top in head section, back section, seat section with perinea cut and split leg section
4. Should have accessory rails on both sides to hold various accessories
5. Should have accessory rails on both sides to hold various accessories
6. Should be suitable for patient load proof atleast 350 kg..
7. Self compensating floor unlock device
8. Emergency override floor unlock device
9. The mattress should be a pressure management pad mattress with at least 80 mm thickness
10. Should have width extender of at least 5 cm width on either side for obesity procedures
11. Should have a minimum height of 600mm or lesser and should have a height adjustment of at least 300mm.
12. Side tilt 20 degree or more
13. Back section adjustment 80 degree up and 40 degree down or more

14. Leg section adjustment 15 degree up and 90 degree down or more
15. Trendelenburg adjustment at least 30 degree and reverse trendelenburg at least 60 degree.
16. Extendable head rest 60 degree up and 90 degree down
17. Tabletop width (w/o side rails) should be more than 500 mm
18. The table length (with extendable head plate) SHOULD BE ATLEAST 2000 MM
19. Table top should be free of exposed screws.
20. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid
21. Should be supplied with the following bariatric accessories
 1. Padded arm boards – 1 pair
 2. Anesthesia screen frame with clamps -1
 3. Restrain strap -1
 4. Shoulder support pair with clamps -1
 5. Lateral side support pair with clamps- 1
 6. Leg crutches with clamps – 2 nos
 7. Split foot rest pair with clamps – 1 set
22. Should have battery back up