



Running Contract Details	
Equipment Name	BINOCULAR MICROSCOPE - VARIOUS TYPE
Running Contract Valid Till	06-09-2020
Tender Ref No	KMSCL/EP/T265/13/2018
Tendered Quantity	250
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Binocular Microscope - Type II <i>Model & Make : LX 500 / N.K. JAIN INSTRUMENTS PVT LTD.</i>	38320.5 Incl.GST :18%	2682.44	41002.94
		38320.5	2682.44	41002.94

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Binocular Microscope - Type II				
Labour	1,770.00	1,770.00	1,770.00	
Comprehensive	4,130.00	4,130.00	4,130.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Binocular Microscope - Type II

II. BINOCULAR MICROSCOPE – Type II

1. Body - Inter changeable, inclined Binocular body, 360 degree rotatable head, field no. 22
2. Should have revolving nose piece with minimum 4 positions.
3. Eyepieces - Highest quality 10 X wide field anti fungus field eyepiece, FOV 22mm.
4. Objectives - Parfocal, antifungus coated 4x, 10x, 40x and 100x (oil immersion) with full plan achromatic correction for FOV of 22 mm.
5. Optical system - Infinity corrected
6. Stage - Horizontal mechanical stage preferably 100 x 140 mm with fine vernier graduations designed with convenient coaxial adjustment for slide manipulation preferably through 30 x 70 mm.
7. Sub stage - Abbe condenser N.A 1.25 focusable, continuously variable iris diaphragm
8. Illuminator – Kohler illumination with 30 W Halogen illumination/LED –with battery backup.
9. Finish - A durable textured acid resistant finish.
10. Should provide dust cover, immersion oil & 1 spare lamp(in case of halogen).
11. Should work with input 200 to 240Vac 50 Hz supply.
12. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.
13. Additional polarising unit shall be quoted separately in BOQ

Miscellaneous Details					
#	Name of the Item	Basic Price(Rs)	GST(Rs)	Cust.Tax(Rs)	Misc.Charges(Rs)
1	polarising Unit	1980.00	237.60	0.00	0.00