

Running Contract Details	
Equipment Name	Binocular Microscope - Type III
Running Contract Valid Till	21-08-2021
Tender Ref No	KMSCL/EP/T321/13C/2019
Tendered Quantity	800
Supplier Name	M/s Lawrence & Mayo (I) Pvt Ltd
GST No	32AAACL1880A1Z6
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Door No. 62/4040 Iyyattil Junction Chittoor Road Ernakulam	Contact Person	Mr. C. Pramod
	Phone	4842368654
	Mobile No	9539721889
	Email	pramod@lawrenceandmayo.co.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Binocular Microscope - Type III <i>Model & Make : DM-R L (DM-R Series) / Debro Engineers, Haryana</i>	13570 Incl.GST :18%	949.9	14519.9
		13570	949.9	14519.9

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Binocular Microscope - Type III				
Labour	500.00	500.00	500.00	
Comprehensive	1,000.00	1,100.00	1,200.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Binocular Microscope - Type III

1. Inter changeable, inclined Binocular body, 360 degree rotatable head.
2. Eyepieces - Highest quality 10 X wide field anti fungus field eyepiece.
3. Objectives- Parfocal, antifungus coated 4x, 10x, 40x and 100x (oil immersion) with semiplan achromatic correction for FOV of 18 mm.
4. Sub stage - Abbe condenser N.A 1.25 focusable, continuously variable iris diaphragm
5. Illuminator - Built-in LED/ Halogen light source with white light.
6. Finish - A durable textured acid resistant finish.
7. Should provide dust cover and immersion oil.
8. Should have Electrical safety certification
9. Should work with input 200 to 240Vac 50 Hz supply