



Running Contract Details	
Equipment Name	Bipolar cautery - ophthal
Running Contract Valid Till	26-11-2020
Tender Ref No	KMSCL/EP/T282/14/2018(R)
Tendered Quantity	50
Supplier Name	M/s Nethra Instrument
GST No	33AAFFN1864Q1ZU
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
MIG 134 - 8th Street Mogappair Eri Scheme Chennai - 600037	Contact Person	Mr.R Vasimalai, Mr. R Anandhan
	Phone	044-26530672
	Mobile No	9840033073, 9847063262
	Email	nethra.ints@gmail.com, nethravaasu@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Bipolar cautery - ophthal <i>Model & Make : WFB-02 / Icon Medicare</i>	25760 Incl.GST :12%	1899.8	27659.8
		25760	1899.8	27659.8

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Bipolar cautery - ophthal				
Labour	4,500.00	4,500.00	4,500.00	4,500.00
Comprehens ive	6,000.00	6,000.00	6,000.00	6,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Bipolar cautery - ophthal

1. Should have use solid state technology.
2. Should have LED indicator for power output.
3. Should have foot control.
4. Should have provision to connect wide range of bipolar forceps and hemostatic erasers to facilitate most ophthalmic surgical procedures.
5. Should be supplied with two sets of reusable, autoclavable cord for bipolar forceps and hemostatic erasers.
6. Should be supplied with one standard ophthalmic bipolar forceps.
7. Should be supplied with 2 bipolar eraser with 18 G tip.
8. Should work with input 200 to 240Vac 50 Hz supply.