

Running Contract Details	
Equipment Name	Bone Densitometer (DEXA Machine)
Running Contract Valid Till	20-12-2020
Tender Ref No	KMSCL/EP/T288/628/2018(R)
Tendered Quantity	15
Supplier Name	M/s Wipro GE Healthcare Pvt Ltd
GST No	29AAACW1685J1ZW
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
No. 4 Kadugodi Industrial Area Whitefield Bangalore 560067	Contact Person	Mr. John Tenny
	Phone	080-28452923
	Mobile No	9895895006
	Email	tenny.john@ge.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Bone Densitometer (DEXA Machine)</b> <i>Model &amp; Make : prodigy advance bone densitometer / GE Health Care/ Wipro GE Health Care</i>	2500000 Incl.GST :0%	206500	2706500
		<b>2500000</b>	<b>206500</b>	<b>2706500</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Bone Densitometer (DEXA Machine)</b>							
Labour	90,000.00	94,500.00	99,500.00	1,05,000.00	1,10,250.00	1,16,000.00	1,22,000.00
Comprehensive	2,09,475.00	2,19,949.00	2,30,946.00	2,42,493.00	2,54,618.00	2,67,349.00	2,80,717.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Bone Densitometer (DEXA Machine)**

1. X-ray Source: Constant Potential Source/ Switched Pulse Dual Energy
2. Detector System : Multi Element / Direct Digital Detectors - solid state with fast pulse counting technology
3. Scan Time: A/P Spine  $\leq$  30 Secs; Femur  $\leq$  30 Secs
4. Scanning Method- Fan Beam
5. Calibration: Automatic calibration Technique for test Programme & quality control
6. BMD Precision : Better Than 1%
7. Patient Position : Cross Hair Laser Light
8. Scan region - 190cmx60cms or more for total body
9. Patient Weight Limits: 150 Kgs minimum
10. Computer System:
  - a) Workable with most Advanced configuration
  - b) Hard Disk Minimum - 500GB
  - c) RAM : Minimum 4GB
  - d) CD /DVD drive
  - e) Monitor: At least 17" Color monitor
  - f) Printers: Laser Printer
11. Software required for the following:
  - a) Total Body with Body Composition
  - b) Comparison to previous Scan
  - c) Lateral spine BMD
  - d) AP Spine

e) Vertebral Assessment ( AP & Lateral Views)

f) Fore Arm

g) Dual Femur

h) Orthopedic Hip Analysis

i) Visceral Fat Analysis

j) Pediatric software ( Spine & Total Body for age group 5-19)

k) DICOM Compatibility

l) Composer

12. Online UPS: 2KVA with 30 mins Backup

13. AERB Type approval: Type approval shall be provided for the model quoted

14. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.