

Running Contract Details	
Equipment Name	Brain Evoked Response audiometry(BERA)
Running Contract Valid Till	16-01-2020
Tender Ref No	KMSCL/EP/T248/1003/2017(R)
Tendered Quantity	28
Supplier Name	M/s Medilife Technologies
GST No	07AAJFM2355B1ZL
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
1st Floor Plot No. 2 Vikas Plaza LSC Kalkaji New Delhi-110 019	Contact Person	Mr. Sanjay Sharma
	Phone	01140588607
	Mobile No	9810524302
	Email	info@medilife.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Brain Evoked Response audiometry(BERA)</b> <i>Model &amp; Make : Neuro Audio/ABR with ASSR / Neurosof Ltd.Russia</i>	653100 Incl.GST :5%	51377.2	704477.2
		<b>653100</b>	<b>51377.2</b>	<b>704477.2</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Brain Evoked Response audiometry(BERA)</b>							
Labour	10,620.00	10,620.00	10,620.00	10,620.00	10,620.00	10,620.00	10,620.00
Comprehens ive	14,160.00	14,160.00	14,160.00	14,160.00	14,160.00	14,160.00	14,160.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Brain Evoked Response audiometry(BERA)**

1. Should have ABR, MLR, LLR, ASSR
2. Transducers: insert phone –Ear Tone 3 A, Head phone, Bone oscillator
3. Should have minimum 2 channels.
4. Polarity: Rarefaction, condensation, alternating
5. Should have Intensity: 0-99 nHL or above
6. Stimulus: Clicks, Tone burst
7. Should have user friendly repetition rate
8. Should have Notch filter 50 or 60 Hz
9. Should have Ipsi and contra noise masking
10. Should be Portable & built in isolation and shielding and test should be carry out in NICU or any locations in hospital
11. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid
12. ASSR: up to 4 stimulus frequency per ear, intensity 0 to 125 dp SPL, frequency range 250 Hz to 8 KHz, users defined modulation, simultaneous testing of both ears
13. Upgradable to : Diagnostic DPOA,TEOAE – Adjustable pass criteria, frequency dominant spectrum vies
14. Ear probe tips – 2 packets (assorted)
15. Reusable cup electrodes – 6 nos
16. T 20 conduction gel – 3 nos
17. Nuprep gel 114 gm - 6 nos
18. Desk Top Computer
  - a. Processor: Intel core i3

- b. RAM : 4 GB
- c. Hard Disk: 1TB
- d. CD/ DVD
- e. Monitor : LED/LCD 18.5 inch or above
- f. Resolution : Minimum display with 1366X768
- g. Operating System: Genuine window 8.1 SL or above
- h. Port: Minimum 4 USB
- i. Should have antivirus software

19. Laser Printer

- a. ISI/ISO certified auto duplex printer
- b. Resolution: Minimum 600 X 600 dpi
- c. Print Speed: 18 PPM
- d. Port: 2 USB or higher

20. Should be provided with table/trolley (Wooden/Ply/ MDF) to keep BERA machine, computer, UPS and printer

21. A single UPS of suitable capacity with 30 minutes back up for BERA, Desktop and Printer shall be provided