



Running Contract Details	
Equipment Name	CPM (Upper Limb)
Running Contract Valid Till	04-04-2021
Tender Ref No	KMSCL/EP/T315/24B/2019(R)
Tendered Quantity	25
Supplier Name	M/s Bharat Bio Products
GST No	33AHLPB1264D1ZY
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
No 188 1st Main Road Nehru nagar OMR kottivakam Chennai TamilNadu 600096	Contact Person	BALASUBRAMANIAN
	Phone	
	Mobile No	9444052532
	Email	bharatbioproducts@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>CPM (Upper Limb)</b> <i>Model &amp; Make : Artromot-S3 / DJO India Healthcare Pvt Ltd</i>	627200 Incl.GST :12%	46256	673456
		<b>627200</b>	<b>46256</b>	<b>673456</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>CPM (Upper Limb)</b>							
Labour	5,000.00	5,000.00	5,000.00	6,000.00	6,000.00	6,000.00	6,000.00
Comprehensive	25,000.00	25,000.00	25,000.00	26,000.00	26,000.00	26,000.00	26,000.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :CPM (Upper Limb)**

1. Should provide anatomically correct alignment
2. Should give normal anatomical range of motion
3. Unit should stop when patient resistance increase the target set value
4. Should have storage card for storing programmed values
5. Facility to use right or left extremity with ease
6. Provision for elbow and shoulder CPM
7. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.