



Running Contract Details	
Equipment Name	Cell saver
Running Contract Valid Till	21-09-2020
Tender Ref No	KMSCL/EP/T271/1163/2018(R)
Tendered Quantity	5
Supplier Name	M/s Haemolab Equipments
GST No	32AAHFH4793A1ZN
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
No.64/3652 (Old 32/2605 â€“ D) - Nilavareth Lane Azad Road - Kaloor Cochin - 682017	Contact Person	Achutha Prakash AJ
	Phone	0484 6999966
	Mobile No	90371 77781, 09387645876
	Email	prakashachutha@gmail.com, sales@haemolab.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Cell saver <i>Model & Make : C.A.T.S Plus / Fresenius Kabi, Germany</i>	1456000 Incl.GST :12%	107380	1563380
		1456000	107380	1563380

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Cell saver							
Labour	0.00	0.00	0.00	25,000.00	25,000.00	26,000.00	26,000.00
Comprehen sive	70,000.00	70,000.00	70,000.00	75,000.00	75,000.00	80,000.00	80,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Cell saver

1. The equipment should be compact, lightweight, having castor wheels for easy movement in operation theatre and should work on 220V/AC.
2. It should work in fully automatic mode as well as in manual mode.
3. The equipment should have in-built vacuum pump to suck mode.
4. The centrifuge should be adjustable from 1500 to 5400 RPM and pump flow from 25 to 1000 ml per minute.
5. It should have display to show all information during the operation as pump speed, centrifuge speed, and alert messages.
6. The equipment should be able to separate lost blood, anticoagulant, filter, store, concentrate, and wash.
7. Beside the RBC separation and washing, it should be able to sequester plasma and platelet from salvaged blood in separate bags.
8. Five years warranty including all components should be given.
9. Should have European CE with 4 digit notifying body number / USFDA approved certificate to be provided.