



Running Contract Details	
Equipment Name	Centrifuge - 8 Wells
Running Contract Valid Till	22-03-2021
Tender Ref No	KMSCL/EP/T304/21C/2018(R)
Tendered Quantity	500
Supplier Name	M/s Kochin Inlab Equipments India Pvt Ltd
GST No	32AADCK3263F1ZX
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Door #28/2391-D Vihaan - Chilavannur Kadavanthra P.O Kochi-682020.	Contact Person	Manoj
	Phone	
	Mobile No	9447790938
	Email	manoj.inlab@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Centrifuge - 8 Wells <i>Model & Make : R-8C PLUS (Centrifuge) S 8-15 (Rotor) / REMI ELEKTROTECHNIK LTD</i>	29736 Incl.GST :18%	2081.52	31817.52
		29736	2081.52	31817.52

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Centrifuge - 8 Wells				
Labour	1,300.00	1,430.00	1,580.00	
Comprehensive	2,600.00	2,860.00	3,160.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Centrifuge - 8 Wells

1. Should have a maximum speed of 5000 RPM with stepless regulator
2. Should be supplied with safety lid and lock.
3. Should have digital speedometer and timer.
4. Should have imbalance detector and automatic cutoff.
5. Standard mark on quality and safety is Compulsory.
6. Should work on 200-240Vac 50Hz power supply.
7. Should have swing out rotor of size 8*15ml