



Running Contract Details	
Equipment Name	Composite light Cure Unit
Running Contract Valid Till	22-09-2021
Tender Ref No	KMSCL/EP/T328/400D/2019(R)
Tendered Quantity	500
Supplier Name	M/s X-Ray Medico Surgical Engineering Ancillary
GST No	33AAAFX0226H1ZM
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Old No.45 New No.64 Third Main Road Kasthurba Nagar Adyar Chennai-600020	Contact Person	N. Ramanathan
	Phone	044 24410074
	Mobile No	
	Email	xmse6445@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Composite light Cure Unit</b> <i>Model &amp; Make : Wood pecker(LED-D) / Ms Dental Surgicals</i>	29680 Incl.GST :12%	2188.9	31868.9
		<b>29680</b>	<b>2188.9</b>	<b>31868.9</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 <sup>rd</sup> Year	4 <sup>th</sup> Year	5 <sup>th</sup> Year	
<b>Composite light Cure Unit</b>				
Labour	0.00	0.00	0.00	
Comprehen sive	0.00	0.00	0.00	

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Composite light Cure Unit**

1. Should have a LED lamp and should have a calibrated automatic timer for selection of cure time from 10 to 60 seconds. The cost of replacement of LED has to be included in the warranty as well as CAMC offered.
2. It should be a Composite Visible Light Cure Unit
3. Should have a light intensity range of 850 - 1000 mw/cm square or more.
4. The hand piece should have on/off trigger and audible tones in every 10 sec.
5. Should supplied with autoclavable fiber optic probe with full 360° rotation & it should be a 60 degree angle probe
6. Should be separate unit not attached to Dental unit
7. Should provide an Eye protection shield