

Running Contract Details	
Equipment Name	Computerised interferential therapy unit with TENS
Running Contract Valid Till	17-01-2021
Tender Ref No	KMSCL/EP/T298/274/2018(R)
Tendered Quantity	224
Supplier Name	M/s HMS Medical Systems
GST No	33AAEFH3571K1ZD
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
NO 9 1ST CROSS STREET NEW COLONY PORUR CHENNAI	Contact Person	V.CHINNASAMY
	Phone	24760756
	Mobile No	9444073633
	Email	sales@hms.co.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Computerised interferential therapy unit with TENS <i>Model & Make : UNISTIM / HMS MEDICAL SYSTEMS</i>	39144 Incl.GST :12%	2886.87	42030.87
		39144	2886.87	42030.87

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Computerised interferential therapy unit with TENS				
Labour	700.00	875.00		1,050.00
Comprehensive	1,050.00	1,400.00		1,750.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Computerised interferential therapy unit with TENS

1. Should have 0-29 programmes.
2. Should have 2 separate output channels – Channel 1 & Channel II.
3. Power consumption shall be 46 watts.
4. Output current shall be 0-60mA
5. Output frequency shall be 4000Hz – 4250Hz for Channel 1 depending on the base and spectrum control & 4000Hz (Constant) for Channel II.
6. Interference frequency shall be 0 Hz – 100 Hz for Base & 0 Hz – 150 Hz for Spectrum
7. Should have Front panel green Bar graph indicator for interference frequency indicator.
8. Should work with input 200 to 240Vac 50 Hz supply.
9. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.

Muscle Stimulator

1. Micro controller based operation
2. Should be light weight, compact and user friendly
3. Should have galvanic, interrupted galvanic, Faradic and surged faradic modes
4. TENS with continuous and pulsed modes
5. LCD display
6. Output 0 to 120 V
7. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.
8. The bidder shall offer a Inferential Therapy unit with TENS and muscle stimulator integrated to a single machine