

Running Contract Details	
Equipment Name	Crash Cart
Running Contract Valid Till	07-11-2020
Tender Ref No	KMSCL/EP/T275/286/2018
Tendered Quantity	3000
Supplier Name	M/s Carevel Medical Systems Pvt Ltd
GST No	09AABCC6250C1Z1
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
S-35-36 Site- IV Suhibabad Industrial Area Ghaziabad- 201 010	Contact Person	Mr. Gyanendra Rajvanshi
	Phone	0120-4292637
	Mobile No	9599679916
	Email	rkurup@carevelglobal.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Crash Cart</b> <i>Model &amp; Make : C-4502 / Carevel Medical Systems Pvt Ltd</i>	19800.01 Incl.GST :18%	1386	21186.01
		<b>19800.01</b>	<b>1386</b>	<b>21186.01</b>

### Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Crash Cart**

1. Overall size shall be more than 900mm L x 500mm W x 1500mm H.
2. The crash cart should be made of 25.4mmx18G Stainless steel tubular frame work.
3. All metal parts shall be SS304 Grade.
4. Should have push handle on one side.
5. Should have S.S. shelves, six colored removable bins & two polystyrene lockable storage units with three drawers each.
6. Facility to carry ECG Monitors, Defibrillators etc on open areas at top centre and bottom shelves.
7. Should have Stainless steel saline rod fixed with.
8. Two accessory mounting brackets to mount accessories anywhere without the need of pre-threaded holes.
9. Crash cart should be mounted on 12.5 cms dia non-rusting swiveling dual castor wheels. Two having locking arrangement.
10. Oxygen cylinder stand –Stainless steel, on one side.