



Running Contract Details	
Equipment Name	Cryosurgery Unit
Running Contract Valid Till	17-12-2020
Tender Ref No	KMSCL/EP/T288/26/2018(R)
Tendered Quantity	100
Supplier Name	M/s Inskon Medical Solutions Private Limited
GST No	32AACCI7233A1Z8
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
98C Manthrakkal Junction Thaikkattukara Post Aluva Kochi- 683106	Contact Person	ajesh
	Phone	
	Mobile No	9744160009
	Email	sales@inscon.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Cryosurgery Unit <i>Model & Make : Wallach LL100/LLC02 / Cooper Surgicals</i>	386400 Incl.GST :12%	28497	414897
		386400	28497	414897

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Cryosurgery Unit				
Labour	10,000.00	11,000.00	12,000.00	
Comprehensive	25,000.00	30,000.00	35,000.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Cryosurgery Unit

1. Should be stand-alone unit consist of a tank, a pressure regulator, and a probe attached by tubing to the tank.
2. Should be a fully mobile cryo-surgical system with a wide array of interchangeable probes designed for the use of gynaecologic surgical procedures. The interchangeable probes must include both different sizes for the cervix besides flat one for vaginal lesions.
3. Either Nitrous oxide or carbon dioxide can be used as refrigerant.
4. Units should support various probes and tips.
5. Nitrous based unit should have scavenging ability.
6. Should have adjustable freezing temperatures and pressures through a regulation system.
7. Should have non-electric defrosting system.
8. All cryo probes must be autoclavable.
9. Operating pressure 450 to 800 psi.
10. Operating temperature -70 to -10°C for Carbon dioxide and -90 to -25°C for Nitrous oxide.
11. Should be supplied with triggers/connection for N₂O or CO₂.
12. Should be supplied with rolling cart.
13. Should be supplied with unfilled cylinder for N₂O or CO₂.

Should be supplied with all kinds of probes required for gynecology