

Running Contract Details	
Equipment Name	Deep Freezer (-20C)
Running Contract Valid Till	15-02-2021
Tender Ref No	KMSCL/EP/T299/170/2018
Tendered Quantity	100
Supplier Name	M/s Labline Instruments
GST No	32AABFL5818N1Z6
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
P B No. 36 Eroor Road Trippunithura-682301	Contact Person	Bini
	Phone	04842776582
	Mobile No	9020304422
	Email	lablinecochin@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Deep Freezer (-20C) <i>Model & Make : vDF-2001 / LABLINE INSTRUMENTS,COCHIN</i>	74340 Incl.GST :18%	5203.8	79543.8
		74340	5203.8	79543.8

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
Rate	3 rd Year	4 th Year	5 th Year
Deep Freezer (-20C)			
Labour	1,250.00	1,250.00	1,250.00
Comprehensive	5,000.00	5,000.00	5,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Deep Freezer (-20C)

1. Capacity : 200 litre
2. Inner shelf : 3 Nos
3. Refrigerant : CFC free
4. Temperature control: Micro-processor controlled digital display with temperature resolution of 0.1 degree C
5. Alarms: high/ low temperature
6. Door closing and locking with key
7. Adjustment: adjustable leveling feet standard (optional casters)
8. Power supply : 210-240V/50-60Hz.
9. Should be a vertical type freezer
10. Inner chamber should be made of plastic/SS and outer should be GI/CRCA powder coated