

Running Contract Details	
Equipment Name	Diathermy Dermatology
Running Contract Valid Till	27-11-2020
Tender Ref No	KMSCL/EP/T275/32/2018
Tendered Quantity	100
Supplier Name	M/s Sree Gokulam Healthcare Pvt Ltd
GST No	32AAJCS3502K1ZI
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Dhanya Complex Sastha Temple Road Ernakulam North PO Cochin-682018	Contact Person	Anoop
	Phone	
	Mobile No	9995813333
	Email	sales@gokulamhealthcare.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Diathermy Dermatology <i>Model & Make : Hyfrecator 2000 / Conmed Corporation</i>	90608 Incl.GST :12%	6682.34	97290.34
		90608	6682.34	97290.34

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Diathermy Dermatology							
Labour	5,000.00	5,000.00	5,000.00	5,000.00	5,000.00	5,000.00	5,000.00
Comprehensive	15,000.00	15,000.00	15,000.00	16,000.00	16,000.00	16,000.00	17,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Diathermy Dermatology

1. Should have patient isolation transformer for Complete Safety.
2. Should have Insulated electrode to minimize pain & Skin damage.
3. Intensity should be finely adjusted (linear control).
4. Frequency range 0.20 to 2.93 MHz.
5. RF Power Cut - 70% cut 30% coag
 Coag - 60% coag 40% cut
6. Fulguration mode should be available.
7. Should have a digital display for selected power.
8. Should be supplied with full set of reusable electrodes and should be specified in the bid.
9. Should be light weight and portable.
10. Should operate on 220/230 50 Hz, Ac supply.
11. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.