



Running Contract Details	
Equipment Name	Digital Dental Radiography System
Running Contract Valid Till	02-05-2020
Tender Ref No	KMSCL/EP/T257/640/2017
Tendered Quantity	50
Supplier Name	M/s Villa India
GST No	07AACPN9000D1Z6
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
B-22 Naraina Industrial Area Phase -I New Delhi - 110 028	Contact Person	Mr V L Narayan
	Phone	011-25891801
	Mobile No	9810153050
	Email	welcome@villaindia.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Digital Dental Radiography System <i>Model & Make : Ez Sensor HD / M/s Vatech, Korea</i>	257600 Incl.GST :12%	18998	276598
2	Dental X-ray machine and lead apron <i>Model & Make : Clearay DC Mobile X-Ray / OEM Init</i>	123200 Incl.GST :12%	9086	132286
		380800	28084	408884

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Digital Dental Radiography System							
Labour	10,000.00	10,000.00	20,000.00	20,000.00	20,000.00	20,000.00	20,000.00
Comprehensive	25,000.00	25,000.00	40,000.00	40,000.00	40,000.00	40,000.00	40,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Digital Dental Radiography System

I Sensor Specification

- 1 Pixel Size 20X20mm/ 35µm or better
- 2 Theoretical Resolution 20lp/mm or better
- 3 Dynamic Range 12 Bit
- 4 Thickness 0.19”Max
- 5 Total Cable Length 7 ft. or more
- 6 Interface- Direct Interface with USB 3.0/2.0

II X Ray Specification

- 7 Generator Constant Potential (DC)
 - 8 Focal Spot 0.7mm or better
 - 9 Tube voltage Selectable 60-70KVp or better
 - 10 Current Selectable 4, 7mA or better
 - 11 Exposure time 0.04-2 Sec
 - 12 Arm Length Atleast 80”
 - 13 Line Voltage 220-230Vac
 - 14 Accessories Round 8” cone Stand Model with four legs and Wheels
 - 15 Should Supply a. Lead Apron with Thyroid guard – Light weight
- b. Laptop – Minimum I3 processor , 14” monitor, 1TB HDD, RAM 4GB, DVD writer, network card, 3USB ports, power supply
 - c. Suitable software
 - d. Disposable cover for sensors for atleast 100 patients

16 The offered model of X ray machine should have AERB type approved. Copy of the certificate should be submitted along with the offer form.