



Running Contract Details	
Equipment Name	Digital pH Meter
Running Contract Valid Till	10-02-2022
Tender Ref No	KMSCL/EP/T351/226/2019(R)
Tendered Quantity	20
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Digital pH Meter <i>Model & Make : ECPH27005E With ECFG7350401B / Thermo Scientific Eutech</i>	115640 Incl.GST :18%	8094.8	123734.8
		115640	8094.8	123734.8

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Digital pH Meter							
Labour	2,000.00	2,000.00	2,000.00	2,000.00	2,000.00	2,000.00	2,000.00
Comprehensive	3,000.00	3,000.00	3,000.00	3,500.00	3,500.00	4,000.00	4,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Digital pH Meter

1. Should have up to 5 point calibration with auto-buffer recognition
2. Should be Quick, easy electrode diagnosis with multiple pH slopes and offset display.
3. Should have Non-volatile memory holds up to 500 data points – time and date – stamped for GLP compliance
4. Should calibrate with up to 5 custom pH buffers – use any pH values that are ?1.0 pH unit apart.
5. Should have Cal-due alarm – no more out – dated calibrations
6. Should have auto-logging function for convenient continuous monitoring.
7. Should have password protection for setup and calibration
8. Should have pH Range up to 14 pH
9. Should have Resolution / Accuracy of 0.1 mV ; ± 0.2 mV.
10. Should have Temp.Range from 0.0 to 100⁰C / 32.0 to 212.0⁰F
11. Should have ATC Probe
12. Should have buffer sets each 20ml for all 5 point calibrations. NIST traceable buffer sets shall be provided along with individual certificate
13. Should have power requirements of 230v, 50HZ ac supply
14. Should have output Rs. 232C (via cable)
15. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid