

Running Contract Details	
Equipment Name	Diode Laser (Dermatology)
Running Contract Valid Till	21-10-2020
Tender Ref No	KMSCL/EP/T267/1157/2018(R)
Tendered Quantity	10
Supplier Name	M/s Aakaar Medical Technologies Pvt Ltd
GST No	27AALCA7587Q1ZK
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
A -801 Heritage Plaza telli galli cross road Andheri east Mumbai - 400069	Contact Person	Mr. Anil Kulkarni
	Phone	02226823949
	Mobile No	9324253669
	Email	anilk@coherentindia.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Diode Laser (Dermatology) <i>Model & Make : 808 Cute / Beijing Honkon Technologies Co.Ltd</i>	1310400 Incl.GST :12%	96642	1407042
		1310400	96642	1407042

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Diode Laser (Dermatology)							
Labour	60,000.00	63,000.00	66,150.00	69,457.50	72,930.37	76,576.89	80,405.73
Comprehensive	1,20,000.00	1,26,000.00	1,32,300.00	1,38,915.00	1,45,860.75	1,53,153.78	1,60,811.47

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Diode Laser (Dermatology)

NAME - 808 nm DIODE HAIR REMOVAL SYSTEM

LASER GENERATION POWER - 2000 W

LASER TYPE - SEMI-CONDUCTOR LASER

LASER WAVE LENGTH - 808 nm

SAPPHIRE SPOT SIZE - 12×12 mm

ENERGY DENSITY - 10-120 j/cm² (continuous adjustable)

PULSE WIDTH - 10-1200 ms(continuous adjustable)

REPETITION FREQUENCY - 0.5-10 hz

COOLING WAY - SEALED INNER –CIRCLE WATER COOLING SYSTEM

TREATMENT HEAD TEMPERATURE - 10-4°C

POWER RATING - 2000 W

POWER SUPPLY - AC 220 V,50/60 Hz

1. The laser system to be provided with three additional doctor protective goggles.
2. Should be supplied with UPS of required capacity along with machine for 30 minutes backup.
3. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission/FDA(US)/STQC CB certificate /STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate /test report shall be produced along with the technical bid