

Running Contract Details	
Equipment Name	Electrical suction apparatus
Running Contract Valid Till	22-01-2021
Tender Ref No	KMSCL/EP/T298/1273/2018(R)
Tendered Quantity	16
Supplier Name	M/s Miti Electronics Pvt Ltd
GST No	33AACCM1588Q1ZZ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Old No 43 New No 115 Justice Ramaswami Street Kamaraj Avenue Adyar Chennai	Contact Person	G Vimal Raj
	Phone	24451922
	Mobile No	9840831177
	Email	vimal@mitie.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Electrical suction apparatus</b> <i>Model &amp; Make : DOMINANT FLEX / Medela AG,Switzerland</i>	280000 Incl.GST :12%	20650	300650
		<b>280000</b>	<b>20650</b>	<b>300650</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Electrical suction apparatus</b>							
Labour	0.00	0.00	3,000.00	3,000.00	3,000.00	3,000.00	3,000.00
Comprehensive	0.00	0.00	7,500.00	7,500.00	7,500.00	7,500.00	7,500.00

### Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Electrical suction apparatus**

1. Should deliver high vacuum of - 90Kpa/-85 Kpa, 675 mm Hg
2. Flow rate range: 35 - 60liters/minute
3. Should provide with Piston Cylinder Technology
4. Should supply with Membrane Vacuum Regulator
5. Should be available autoclavable PSU jars of 3-5 litres (2 no's) and lids with over flow Protection device
6. Should supply with Foot Vacuum Regulator
7. Should be available safety jar with over flow protection device
8. Tubings should be made from silicone
9. Should supply with standard rail which is attached in the original company made mobile trolley.
10. Mobile trolley should have four castors with brakes and On/Off switch.
11. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.