

Running Contract Details	
Equipment Name	Electronic Adult Weighing Scale (Platform Type)
Running Contract Valid Till	02-04-2021
Tender Ref No	KMSCL/EP/T315/285/2019
Tendered Quantity	118
Supplier Name	M/s Transworld Scales Services
GST No	32AWCPS0193C1ZX
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
Beach Road Kollam - 691 001	Contact Person	Ravisanker B.S
	Phone	0474 2742381
	Mobile No	9446332381
	Email	services-transklm1@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Electronic Adult Weighing Scale (Platform Type) <i>Model & Make : Phoenix NEPI60 PW / Nitraj Engineering</i>	12249.58 Incl.GST :18%	857.47	13107.05
		12249.58	857.47	13107.05

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Electronic Adult Weighing Scale (Platform Type)

1. Capacity : 160 kg
2. Accuracy : 100 g
3. Platter Size : 350 mm x 300 mm (Tolerance +/- 10%)
4. The scale should be made up of heavy duty. Cast iron structure Platform with powder coated frames.
5. The Electronic Adult Weighing Scale should incorporate following features for user-friendly convenience.
6. Display: LED / LCD : 5 digit with min. height 14 mm.
7. TARE facility with zero function.
8. The Scale should have inbuilt rechargeable battery backup for minimum of 8 hrs.
9. Should operate on mains 220-240Vac, 50 Hz single phase.
10. The Scale should be as per BIS specifications. The scale should have ISI mark.
11. The display stand height shall be 80cm(Tolerance +/- 10%) from the platform.
12. The equipment shall be supplied with valid stamping from Weight and Measures. The stamping shall be done free of cost during the warranty period and the CAMC rates offered shall include the stamping charges.
13. The tenderer shall have valid sales and service license for Weighing Machines from Legal Metrology.