



Running Contract Details	
Equipment Name	Fetal Monitor Model A
Running Contract Valid Till	02-05-2020
Tender Ref No	KMSCL/EP/T259/48A/2017
Tendered Quantity	300
Supplier Name	M/s Sree Gokulam Healthcare Pvt Ltd
GST No	32AAJCS3502K1ZI
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Dhanya Complex Sastha Temple Road Ernakulam North PO Cochin-682018	Contact Person	Anoop
	Phone	
	Mobile No	9995813333
	Email	sales@gokulamhealthcare.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Fetal Monitor Model A <i>Model & Make : JPD-300P / Shenzhen Jumper Medical Equipment Co.Ltd</i>	77000 Incl.GST :12%	5678.75	82678.75
		77000	5678.75	82678.75

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Fetal Monitor Model A				
Labour	3,245.00	3,245.00		3,245.00
Comprehensive	10,030.00	10,525.00		11,210.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Fetal Monitor Model A

1. Should be portable.
2. Should have the measurement of FHR, and Uterine activity.
3. Should be supplied with event marker cable and an event marker key on the control panel.
4. Should have alert indication for probe contact, high/low FHR.
5. Should have a volume control for FHR.
6. Should have large format LED/LCD/TFT display.
7. Should be provided with thermal printer for printing FHR, uterine activity pressure.
8. Thermal paper should be available locally.
9. Should provide alarm limit settings for high, low FHR.
10. Should provide belt for both ultrasound and toco transducer.
11. Should be supplied with SS trolley.
12. Should operate on mains (230v,50Hz).
13. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission/FDA(US)/STQC CB certificate /STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate /test report shall be produced along with the technical bid.
14. Accessories:
 - i. Gel – 4 bottle
 - ii. Thermal paper- 10Nos
 - iii. Voltage Stabiliser
15. Should have built in rechargeable Li –ion battery with backup of at least one hour
16. The paper speed should be 1 cm per minute