



Running Contract Details	
Equipment Name	Fetal Monitor Model B
Running Contract Valid Till	02-05-2020
Tender Ref No	KMSCL/EP/T259/48B/2017
Tendered Quantity	100
Supplier Name	M/s Sree Gokulam Healthcare Pvt Ltd
GST No	32AAJCS3502K1ZI
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Dhanya Complex Sastha Temple Road Ernakulam North PO Cochin-682018	Contact Person	Anoop
	Phone	
	Mobile No	9995813333
	Email	sales@gokulamhealthcare.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Fetal Monitor Model B <i>Model & Make : JPD-300P / Shenzhen Jumper Medical Equipment Co.Ltd</i>	88200 Incl.GST :12%	6504.75	94704.75
		88200	6504.75	94704.75

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Fetal Monitor Model B				
Labour	3,245.00	3,245.00		3,245.00
Comprehensive	11,210.00	11,800.00		12,390.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Fetal Monitor Model B

1. Should have 10" or more high resolution color TFT/LCD Display with Tilttable Screen for better viewing.
2. Should have the facility for dynamic data save.
3. Should display the monitoring information of last 24 hours is essential.
4. Should have special high sensitive watertight probe for better durability.
5. Should have data storage with play back & print facility.
6. Should have the low ultrasound power for the safety of the foetus.
7. Should have automatic foetal movement detection with event marker.
8. Thermal printer with minimum 152MM paper width is essential for broader print outs.
9. Standard configuration should be FHR, TOCO, Foetal Movement.
10. Twin FHR monitoring is essential.
11. Should be portable.
12. Should supply foetal stimulator.
13. Built in rechargeable Li-on battery with back up of atleast one hour.
14. Should have work with input 200 to 240Vac 50 Hz supply.
15. Central monitoring system software upgradability should be possible(optional).
16. Accessories:
 - i. Gel – 4 bottle
 - ii. Thermal paper- 10Nos
 - iii. Voltage Stabiliser
17. Should provide belt for both Ultrasound and toco transducer.
18. Should be supplied with suitable trolley

19. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.

20. The paper speed should be 1 cm per minute.