



Running Contract Details	
Equipment Name	Flexible Ureteroscope
Running Contract Valid Till	20-11-2020
Tender Ref No	KMSCL/EP/T282/270/2018(R)
Tendered Quantity	17
Supplier Name	M/s Olympus Medical Systems India Pvt. Ltd
GST No	06AABCO2131L1ZO
Installation & Delivery Period	4 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Olympus Medical Systems India Pvt. Ltd Ground Floor Tower- C SAS Tower The Medicity Complex Sector â€™ 38 Gurgaon â€™ 122001	Contact Person	Mr. Neelamani
	Phone	1244999191
	Mobile No	9500069243
	Email	endosale.omsi@olympus-ap.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Flexible Ureteroscope <i>Model & Make : URF-P7R / Olympus Corporation</i>	704000.25 Incl.GST :0%	58150.42	762150.67
		704000.25	58150.42	762150.67

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Flexible Ureteroscope							
Labour	5,019.00	5,521.00	6,073.00	6,680.00	7,348.00	8,083.00	8,891.00
Comprehen sive	25,093.00	25,093.00	25,093.00	25,093.00	25,093.00	25,093.00	25,093.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Flexible Ureteroscope

1. Should be a flexible fiberoptic Uretero – Renoscope with Continuous Irrigation.
2. Outer diameter of the shaft should be between 6 Fr. - 10 Fr.
3. Should have an instrument channel between 3 Fr. to 4 Fr.
4. Should have 680mm \pm 20mm working length..
5. Should have a ceramic liner in the distal end of the working channel to protect it from thermal or electro Cautery damage.
6. Should have angle of view between 800- 900 field of view 00 with eyepiece adjustment.
7. Should have minimum angle of deflection 1800 up, 2700 down.
8. Should have active deflection mechanism.
9. Should be supplied with gas sterilization valve.
10. Should be waterproof and fully immersible in solution.
11. Should be adhering to sterilization method with ETO, FO gas.
12. Should be supplied with the following accessories of compatible sizes.
 - a. Grasping Forceps
 - b. Biopsy Forceps
 - c. Pressure compensation cap
 - d. Leakage Tester
 - e. Cleaning Brush
 - f. Carry Case
 - g. Access sheath
13. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid