

Running Contract Details	
Equipment Name	Fully Manual Rotary Microtome
Running Contract Valid Till	12-04-2020
Tender Ref No	KMSCL/EP/T262/626/2017(R)
Tendered Quantity	10
Supplier Name	M/s Bions Medical Systems Pvt Ltd
GST No	32AACCB4877B1Z1
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
28/3085 D âœRohiniâœ Tagore Nagar Ponneth Temple Road Kadavanthra Kochi-682020.	Contact Person	Jitto John
	Phone	
	Mobile No	9567860501
	Email	bions@vsnl.net

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Fully Manual Rotary Microtome <i>Model & Make : RM2125 RTS / Leica Biosystems GmbH, Germany</i>	528876 Incl.GST :18%	37021.32	565897.32
		528876	37021.32	565897.32

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Fully Manual Rotary Microtome							
Labour	15,866.28	16,659.24	17,492.32	18,366.70	19,285.92	20,249.98	21,262.42
Comprehens ive	37,021.32	38,872.74	40,816.20	42,857.60	45,000.48	47,250.74	49,613.10

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Fully Manual Rotary Microtome

1. Ideally located coarse feed wheel for fatigue – free operation.
2. Extremely smooth-running hand wheel preferably with patented spring-based forced balancing system.
3. Safety locking mechanism on the hand wheel.
4. Section thickness setting range-0.5-60Éµm
5. Trimming thickness ranges – 10,50 Éµm
6. Total horizontal specimen feed – 25mm
7. Vertical stroke length – 59mm
8. Specimen retraction.
9. Specimen orientation – horizontal 80 & vertical 80
10. Coarse feed wheel turn direction : user selectable
11. Single – handed operation specimen clamps
12. Facility to use cassette clamps should be available.
13. Precise knife holder lateral adjustments ensures that the entire width of the blade can be utilized.
14. Universal knife holder base for different knife holders(Disposable, Standard & Heavy duty).
15. Disposable blade holder with integrated safety finger protection guard.
16. Accessories to be supplied alongwith the machine :
 - a. Cassette Clamp -1 No.
 - b. Disposable Microtome Blades – 2 packets (50nos/pack)
 - c. Dust Cover, Oil, Manual