



Running Contract Details	
Equipment Name	Fumigation dispensor
Running Contract Valid Till	12-06-2021
Tender Ref No	KMSCL/EP/T322/50/2019
Tendered Quantity	350
Supplier Name	M/s Biomedical Engineering Company
GST No	32AAGFB1151K1ZV
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
39/878 A2 - YMJ West Lane Palarivattom Kochi 682025	Contact Person	Mr. Ajith George Mathew
	Phone	
	Mobile No	9447012546
	Email	mail@behealthcare.com,accounts@behealthcare.com,sales@behealthcare.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Fumigation dispensor <i>Model & Make : Steri-Fog / Faith Innovations</i>	26762.4 Incl.GST :12%	1973.73	28736.13
		26762.4	1973.73	28736.13

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Fumigation dispensor				
Labour	1,338.00	1,472.00		1,619.00
Comprehensive	2,677.00	2,944.00		3,239.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Fumigation dispenser

1. Should have a consistent particle size generation of less than one microns.
2. Should have a maximum through put of 4 meters in a closed room.
3. Should have a vortex type design, non-rotating and non-clogging nozzle assembly.
4. Should have tough and corrosion resistant power head housing.
5. Should have a chemical solution tank of minimum capacity 5 litres.
6. Should have a tough and corrosion resistant tank. The bolts, nuts and washers used inside the tank should be stainless steel / brass/ corrosion resistant material.
7. Should have an easy to clean, detachable and non-corrosive chemical solution tank.
8. Should have primary and secondary filtration for the air intake system.
9. Should be compatible with wide range of disinfectants.
10. Should operate from 200 to 240Vac, 50 Hz input supply.
11. Should have a timer device
12. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid.