



Running Contract Details	
Equipment Name	Haemoglobinometer
Running Contract Valid Till	27-12-2020
Tender Ref No	KMSCL/EP/T269/52/2018
Tendered Quantity	250
Supplier Name	M/s V.S.Nachiyar Associates
GST No	32AAGFSV8029A1ZH
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
T.C 2/725(3) - RAJEEV GANDHI LANE MEDICAL COLLEGE POST TRIVANDRUM - 695 011	Contact Person	Vellaisamy
	Phone	0471 2556678, 0471 4015343
	Mobile No	9447722332 ,9447777678
	Email	vsnassociates@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Haemoglobinometer (Cuvette Type) <i>Model & Make : Hemocue HB 201+ / Hemocue AB</i>	22514.4 Incl.GST :18%	1576.01	24090.41
2	cuvette	27.14 Incl.GST :18%	1.9	29.04
		22541.54	1577.91	24119.45

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Haemoglobinometer				
Labour	1,000.00	1,000.00	1,000.00	1,000.00
Comprehensiv e	1,000.00	1,000.00	1,000.00	1,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Haemoglobinometer

1. Should have a direct read-out on LED display of hemoglobin in g/l or g/dl.
2. Should be ready for immediate use with no calibration required by the user.
3. Should have auto zeroing and switch to standby mode.
4. Simply inserting the cuvette should activate the meter and initiates reading of a test sample.
5. Should work with input 200 to 240Vac 50 Hz supply.
6. 200nos of cuvettes shall be supplied free of cost along with the machine
7. Cost of 2500 cuvettes will be taken for evaluation