



Running Contract Details	
Equipment Name	Head Light for ENT OT
Running Contract Valid Till	27-12-2020
Tender Ref No	KMSCL/EP/T298/54/2018(R)
Tendered Quantity	200
Supplier Name	M/s Trivandrum Surgicals and Lab Supplies
GST No	32AAIFT1010D1Z0
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
T.C.2/1094(7) SREELAKSHMI PAZHAYA ROAD MEDICAL COLLEGE P.O TVM.695015	Contact Person	GEETHA . S
	Phone	0471 - 2555040
	Mobile No	09447340147
	Email	trivandrumsurgicals@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Head Light for ENT OT <i>Model & Make : KD / VMS</i>	39760 Incl.GST :12%	2932.3	42692.3
		39760	2932.3	42692.3

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Head Light for ENT OT				
Labour	2,485.00	2,840.00	3,195.00	
Comprehensive	2,485.00	2,485.00	2,485.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Head Light for ENT OT

1. Should be 5W LED headlight system suitable for ENT Operating Theater. The system should have rechargeable battery with a minimum back up of 4hrs.
2. The reflector should be a multiple coated.
3. Should have head light adjustment side to side and up and down.
4. Should have focusing sleeves for uniform quality illumination.
5. Should have an adjustable and light weight head band with lock.
6. Should have provision to adjust light intensity.
7. Should work with input 200 to 240Vac 50 Hz supply.
8. Should be provided with an adapter for charging the machine.
9. Should be provided with an additional rechargeable battery.