



Running Contract Details	
Equipment Name	Holter Monitor Model B
Running Contract Valid Till	10-01-2021
Tender Ref No	KMSCL/EP/T290/160B/2018
Tendered Quantity	20
Supplier Name	M/s Schiller Healthcare India Pvt. Ltd
GST No	34AADCS5091B1ZN
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Schiller Healthcare India Pvt Ltd 37/1013 2nd Floor Unidec House Fathima Church Road Elamkulam Kochi-682020	Contact Person	Mr.Stijo Mathew , Mr. P.S Vidhyashankar
	Phone	04842203264
	Mobile No	09846706787, 09383620520
	Email	stijo@schillerindia.com, vidyashankar@schillerindia.com, mail@schillerindia.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Holter Monitor Model B (including all the accessories given below) <i>Model & Make : Medilog FD 12 plus with medilog darwing 2 / Schiller</i>	541184 Incl.GST :12%	39912.32	581096.32
2	Recorder & accessories <i>Model & Make : FD 12 Plus</i>	150640 Incl.GST :12%	11109.7	161749.7
3	Software <i>Model & Make : Darwin Analyser</i>	328720 Incl.GST :12%	24243.1	352963.1
4	Computer <i>Model & Make : Dell/HP</i>	61824 Incl.GST :12%	4559.52	66383.52
		1082368	79824.64	1162192.64

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Holter Monitor Model B							
Labour	4,000.00	4,000.00	4,000.00	4,000.00	4,000.00	4,000.00	4,000.00
Comprehensive	16,200.00	16,200.00	16,200.00	16,200.00	16,200.00	16,200.00	16,200.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Holter Monitor Model B

1. System should provide for 24 hours and 7 days of continuous ECG recording and analysis for detecting heart function abnormalities which may otherwise go undetected. The system should support 3 channel recorder.
2. Should be able to record 24 hours and 7 days of ECG waveforms on small Holter recorders.
3. Should automatically detect and quantify different ventricular and supraventricular events, including atrial events (atrial fibrillation, isolated premature, pairs, bigeminy, trigeminy, runs, pauses, bradycardia and tachycardia) and ventricular events (isolatedectopics, prematureectopics, bigeminy, trigeminy, couplets, triplets and runs).
4. Should employ multiple analysis modes, including prospective, paging and superimposition, retrospective modes.
5. Should analyse three leads of ST segments with ST episode reporting and heart rate variability on time or frequency domain.
6. Should provide unlimited normal, abnormal and artifact templates with automatic classification, template matching and ability to merge /unmerge on any template.
7. Should have provision for patient diary entries and ST episodes.
8. Should have QT and pacemaker analysis.
9. Should create custom reports templates with institution's logo.
10. Trend graphs-HR, RR interval, ST.
11. The system should be supplied with a branded (DELL/HP) desktop/laptop with windows operating system. In case of desktop suitable table and UPS shall be provided. The minimum specification is 8th generation, 8GB RAM, 1 TB hard Disk atleast 24 " LED monitor and the accessories.

I. Specification for Recorder

1. Should weight not more than 180grams without battery.
2. Should acquire simultaneous ECG using atleast 5- lead patient cable for a true 3 channel acquisition.
3. Should automatically remove pacing artifacts and annotates the recording with pacing pulses.
4. Should store 24/48 hours and 7 days of ECGs with no data compression.
5. Use AA/AAA alkaline batteries to provide up to 48 hours and 7 days of continuous recording.
6. The recorder should have display for viewing ECG waveform during hook up.
7. Should be water resistant.
8. Should synchronize the recording end time with the recorder time clock.
9. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.