



| Running Contract Details | |
|--------------------------------|-------------------------------------|
| Equipment Name | ICU COT VARIOUS TYPES |
| Running Contract Valid Till | 22-12-2021 |
| Tender Ref No | KMSCL/EP/T341/58C/2019 |
| Tendered Quantity | 500 |
| Supplier Name | M/s Carevel Medical Systems Pvt Ltd |
| GST No | 09AABCC6250C1Z1 |
| Installation & Delivery Period | 8 Week(s) |
| Up-time / PM vist | 95% & 2 Visits per year |
| Warranty period | 2 Years |

| Supplier`s Details | | |
|---|-----------------|--------------------------|
| Address | Contact Details | |
| S-35-36 Site- IV Suhibabad Industrial Area Ghaziabad- 201 010 | Contact Person | Mr. Gyanendra Rajvanshi |
| | Phone | 0120-4292637 |
| | Mobile No | 9599679916 |
| | Email | rkurup@carevelglobal.com |

| Item-wise Price Details | | | | |
|-------------------------|---|---|------------------------------------|-----------------|
| # | Item Details | Unit Rate (Incl.all taxes & charges) | Service Charges (Through KMSCL) | Grand Total |
| 1 | ICU Cot Model A <i>Model & Make : c-1204A CAREVEL MEDICAL SYSTEMS PVT LTD</i> | 33707.88 Incl.GST :18% | 2359.55 | 36067.43 |
| | | 33707.88 | 2359.55 | 36067.43 |

| Annual / Comprehensive Maintenance Charges (Exl.Tax) | | | |
|--|----------------------|----------------------|----------------------|
| Rate | 3 rd Year | 4 th Year | 5 th Year |
| ICU Cot Model A | | | |
| Labour | 0.00 | 0.00 | 0.00 |
| Comprehensive | 350.00 | 450.00 | 550.00 |

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :ICU Cot Model A

1. Should be a height adjustable, manually operated, four section cot with perforated top.
2. The size should be approximately 2250mm length, 1015mm width and an approximate height adjustable from 520 to 820mm with $\pm 10\%$ tolerance
3. Should have a trendelenburg and reverse trendelenburg movements.
4. Should have swiveling non corrosive, fibre / synthetic wheels of 125mm dia. 2 with brake and 2 without brake.
5. Should provide with a detachable and collapsible railing made of stainless steel of minimum 19mm diameter and 18 gauge tubes housed in mild steel tubular frame, with a locking facility for raised position.
6. Should provide with removable heavy duty saline stand made of stainless steel.304 grade and one telescopic saline stand with two hooks. . Should have a provision to fix the telescopic saline stand on all four corners of the bed and also in the middle of the bed on either side
7. Should be provided with mattress of 4 section, 4” thick PU foam of 40 density covered with soft water proof material, bacteriostatic & pillow
8. Should be supplied with mattress size should be approximately 2010 mm x 910 mm x 100 mm with 10% tolerance.
9. All MS components should be 7 dip tank pre-treated and epoxy powder coated with 50-60 microns.
10. ICU bed should be of minimum 130Kg weight bearing capacity.
11. All MS Sheets/tubes must be CRCA
12. SS components wherever used should be of 304 grade.
13. Oxygen Cylinder holder should be provided.
14. The mattress should be supplied as per the specification. Also need to provide pillows
15. The item supplying should have the brand name and model name etched/embossed to the item
16. Should have SS laminated detachable /fixed head and foot boards