



Running Contract Details	
Equipment Name	Indirect Ophthalmoscope Model B
Running Contract Valid Till	12-02-2021
Tender Ref No	KMSCL/EP/T304/61A/2018(R)
Tendered Quantity	54
Supplier Name	M/s Shree Ganesh Enterprises
GST No	33AFKPD9582K1Z1
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
2B HONEY TONE APARTMENT SANKARAPURAM 1ST STREET CHOO LAIMEDU CHENNAI-600094	Contact Person	MR.SRINIVASAN
	Phone	
	Mobile No	9444161083
	Email	shreegan_02@yahoo.co.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Indirect Ophthalmoscope Model B</b> <i>Model &amp; Make : KEELER VANTAGE PLUS / KEELER, UK</i>	149500 Incl.GST :12%	11025.62	160525.62
		<b>149500</b>	<b>11025.62</b>	<b>160525.62</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Indirect Ophthalmoscope Model B</b>							
Labour	4,500.00	4,500.00	5,000.00	5,500.00	6,000.00	6,500.00	7,000.00
Comprehensive	7,500.00	8,500.00	9,500.00	10,500.00	11,500.00	12,500.00	13,500.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Indirect Ophthalmoscope Model B**

1. Should have a adjustable, light weight, soft cushioning headband.
2. Should have unique Synchronized adjustment of Convergence and Parallax.
3. Illumination beam should be adjustable separately from viewing column.
4. Should have inter-pupillary distance from 48-74mm or better
5. Should have integrated flip-up adjustment in the headband.
6. Should have headband rheostat control knob to increase and decrease the light intensity.
7. Should have apertures like pinhole, small spot, large spot and diffuser.
8. Should have integrated filters like cobalt blue and red-free.
9. Should have LED illumination system.
10. Should have rechargeable battery with charging transformer
11. Should provide 1no of 20 D lens.
12. Should be supplied with three pencils, fundus chart, Scleral depressor and carrying case
13. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.