



Running Contract Details	
Equipment Name	Indirect Ophthalmoscope
Running Contract Valid Till	26-07-2020
Tender Ref No	KMSCL/EP/T264/61/2017
Tendered Quantity	50
Supplier Name	M/s Appasamy Associates
GST No	33AASCA3660R1Z0
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
# 20 - SBI Officers Colony 1st Street Arumbakkam Chennai-600106	Contact Person	B Radhakrishnan
	Phone	04430101400, 30101401
	Mobile No	09349211170
	Email	info@appasamy.com, appasamy.thrissur@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Indirect Ophthalmoscope</b> <i>Model &amp; Make : AAIO / Appasamy Associates</i>	56000 Incl.GST :12%	4130	60130
2	<b>20D lens</b> <i>Model &amp; Make : AAIO</i>	16800 Incl.GST :12%	1239	18039
3	<b>90D lens</b> <i>Model &amp; Make : AAIO</i>	22400 Incl.GST :12%	1652	24052
		<b>95200</b>	<b>7021</b>	<b>102221</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
Rate	3 <sup>rd</sup> Year	4 <sup>th</sup> Year	5 <sup>th</sup> Year
<b>Indirect Ophthalmoscope</b>			
Labour	590.00	590.00	590.00
Comprehens ive	1,180.00	1,770.00	2,360.00

## **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

### **Equipment :Indirect Ophthalmoscope**

Indirect Ophthalmoscope with 20D lens

1. Should have a adjustable, light weight, soft cushioning headband
2. Should have unique Synchronized adjustment of convergence and parallax
3. Should have increased pupillary distance from 48-74mm or better
4. Should have integrated flip-up adjustment in the headband
5. Should have headband rheostat control knob to increase and decrease the light intensity
6. Should have apertures like pinhole, small spot, large spot and diffuser
7. Should have integrated filters like cobalt blue, red free and yellow/amber
8. Should have illumination mirror which can be moved independently of viewing assemblies, spot size, illumination light axis and observation light axis
9. Should have LED illumination system
10. Should have mobile battery pack with charging transformer
11. Should provide 20D lens -1no, Scleral Depressor and Fundus chart
12. Should operate from 200 to 240v, 50/60Hz power supply
13. Should have safety certificate from a competent authority CE/FDA (US). Other equivalent certificates will not be accepted. Should have valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/test report shall be produced along with the technical bid.

Indirect Ophthalmoscope with 90D lens

1. Should have a adjustable, light weight, soft cushioning headband
2. Should have unique Synchronized adjustment of convergence and parallax
3. Should have increased papillary distance from 46-74mm
4. Should have integrated flip-up adjustment in the headband
5. Should have headband rheostat control knob to increase and decrease the light intensity
6. Should have apertures like pinhole, small spot, large spot and diffuser
7. Should have integrated filters like cobalt blue, red free and yellow
8. Should have illumination mirror which can be moved independently of viewing assemblies, spot size, illumination light axis and observation light axis
9. Should have LED illumination system
10. Should have mobile battery pack with charging transformer
11. Should provide 2nos of Spare bulb, 90D lens -1no, Scleral Depressor and Fundus chart.
12. Should operate from 200 to 240v, 50/60Hz power supply
13. Should have safety certificate from a competent authority CE/FDA (US). Other equivalent certificates will not be accepted. Should have valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/test report shall be produced along with the technical bid.