



Running Contract Details	
Equipment Name	Inoculation Hood
Running Contract Valid Till	18-07-2020
Tender Ref No	KMSCL/EP/T265/509/2017
Tendered Quantity	15
Supplier Name	M/s Labline Instruments
GST No	32AABFL5818N1Z6
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
P B No. 36 Eroor Road Trippunithura-682301	Contact Person	Bini
	Phone	04842776582
	Mobile No	9020304422
	Email	lablinecochin@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Inoculation Hood <i>Model & Make : MODEL LIH-I / LABLINE INSTRUMENTS</i>	13983 Incl.GST :18%	978.81	14961.81
		13983	978.81	14961.81

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Inoculation Hood				
Labour	1,180.00	1,180.00		1,180.00
Comprehensive	2,360.00	2,360.00		2,360.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Inoculation Hood

1. Should be of size approx. 2.5 x 2 x 2 ft (lxbxh)
2. Should be made of powder coated steel.
3. Front should be of tempered glass or polycarbonate.
4. Should have illumination with 2 ft fluorescent tube.
5. Should have an inbuilt UV germicidal lamp.
6. Should have an exhaust system.
7. The hood should be ducted out by the supplier, if required.
8. Should have safety certificate from a competent authority CE/FDA (US). Other equivalent certificates will not be accepted. Should have valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/test report shall be produced along with the technical bid