



Running Contract Details	
Equipment Name	Intra Oral digital radiography system
Running Contract Valid Till	12-06-2021
Tender Ref No	KMSCL/EP/T324/640B/2019(R)
Tendered Quantity	50
Supplier Name	M/s Villa India
GST No	07AACPN9000D1Z6
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
B-22 Naraina Industrial Area Phase -I New Delhi - 110 028	Contact Person	Mr V L Narayan
	Phone	011-25891801
	Mobile No	9810153050
	Email	welcome@villaindia.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Intra Oral digital radiography system <i>Model & Make : Vistascan Mini Plus / Durr Dental , Germany</i>	532000 Incl.GST :12%	39235	571235
2	Desktop computer with all accessories	94400 Incl.GST :18%	6608	101008
		626400	45843	672243

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Intra Oral digital radiography system							
Labour	11,100.00	12,210.00	13,431.00	14,774.00	16,251.00	17,876.00	19,665.00
Comprehensive	55,500.00	61,050.00	67,155.00	73,870.00	81,257.00	89,382.00	98,321.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Intra Oral digital radiography system

Equipment Name: Intra Oral Digital Radiography System with accessories

1. Should be able to use along with the existing X-ray system in the department.
2. Should have the following component
 - i. Imaging plates (size 0-4).
 - ii. Scanner for the imaging plate.
 - iii. Desktop computer.
 - iv. Accessories : Imaging plates size 2(6nos), size0(2nos), size4(2nos)
3. Specification of Imaging Plate
 - i. Should be flexible and thin as film for simple positioning.
 - ii. Cable- free system
 - iii. Sizes: 5 sizes should be available such as: size 0: 2x3cm, size 1: 2x4 cm, size 3: 2.7x5.4cm, size 4: 5.7x7.5cm.
 - iv. Broader exposure time frame & wide grey scale.
4. Specification for scanner:
 - i. Should have input/output buffer; with its cassette transport system, Table top type.
 - ii. Should read out and deletes up to 6 images or more in one step and makes the plates immediately available again.
 - iii. Scanning time should be less than 1 minute
5. Specification for computer

Technical specification

1. Processor Intel i5 or its higher version

2. RAM 8 GB with 32 GB expandability
3. Hard disk 1TB
4. Operating system Windows 10 professional
5. Monitor 20" LED/TFT colour monitor
6. Keyboard & mouse 104 keys, optional mouse
7. Ports 4 USB ports or more
8. DVD ROM Drive 8x or better DVD RW drive
9. Antivirus software Microsoft security essentials
10. Network facility Should be available
11. Extra software Abode reader

6. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.