



| Running Contract Details | |
|--------------------------------|--|
| Equipment Name | Intubating Video Laryngoscope with Macintosh Blade |
| Running Contract Valid Till | 07-04-2021 |
| Tender Ref No | KMSCL/EP/T311/1272/2019(R) |
| Tendered Quantity | 16 |
| Supplier Name | M/s Lakshmi Surgicals |
| GST No | 32AACFL2603A1ZB |
| Installation & Delivery Period | 8 Week(s) |
| Up-time / PM vist | 95% & 4 Visits per year |
| Warranty period | 3 Years |

| Supplier`s Details | | |
|--|-----------------|--------------------------------|
| Address | Contact Details | |
| 203XIX -City Towers Bank Jn -Market Road Aluva -Ernakulam - 683101 Kerala | Contact Person | Mr. Viji Vijayan |
| | Phone | |
| | Mobile No | 99466 02090 |
| | Email | lakshmisurgicals2005@gmail.com |

| Item-wise Price Details | | | | |
|-------------------------|--|---|------------------------------------|-------------|
| # | Item Details | Unit Rate (Incl.all taxes & charges) | Service Charges (Through KMSCL) | Grand Total |
| 1 | Intubating Video Laryngoscope with Macintosh Blade C-MACÂ® Monitor for CMOS Endoscopes, screen size 7" - 8403ZXK <i>Model & Make : Karl Storz Intubating Video Laryngoscope with Macintosh Blade / Karl Storz GmbH & Co</i> | 711970.7 Incl.GST :18% | 49837.95 | 761808.65 |
| 2 | C-MACÂ® Connecting Cable, with C-MACÂ® system <i>Model & Make : 8403X</i> | 75984.92 Incl.GST :9% | 5318.94 | 81303.86 |
| 3 | C-MACÂ® Video Laryngoscope MAC #3, CMOS technology, with MACINTOSH laryngoscope blade, size 3, <i>Model & Make : 8403AXC</i> | 410728.64 Incl.GST :6% | 30291.24 | 441019.88 |
| 4 | C-MACÂ® Video Laryngoscope MAC #4, CMOS technology, with MACINTOSH laryngoscope blade, size 4, <i>Model & Make : 8403BXC</i> | 410728.64 Incl.GST :6% | 30291.24 | 441019.88 |

| Item-wise Price Details | | | | |
|-------------------------|--|---------------------------|------------------|-------------------|
| 5 | C-MACÂ® Video Laryngoscope MAC #2, CMOS technology, with MACINTOSH laryngoscope blade, size 2, <i>Model & Make : 8403KXC</i> | 401989.28 Incl.GST :6% | 29646.71 | 431635.99 |
| 6 | Protective Bag, blue, for C-MAC Â®system, <i>Model & Make : 8403YD</i> | 54632.82 Incl.GST :9% | 3824.3 | 58457.12 |
| 7 | IV-Stand for C-MAC monitor <i>Model & Make : 8401YA</i> | 83988.86 Incl.GST :9% | 5879.22 | 89868.08 |
| 8 | MAGILL Forceps, modified by BOEDEKER, length 25 cm <i>Model & Make : 809125</i> | 27061.44 Incl.GST :6% | 1995.78 | 29057.22 |
| 9 | MAGILL Forceps, for children, modified by BOEDEKER, length 20 cm, <i>Model & Make : 809120</i> | 25268.32 Incl.GST :6% | 1863.54 | 27131.86 |
| | | 2202353.62 | 158948.91 | 2361302.53 |

| Annual / Comprehensive Maintenance Charges (Exl.Tax) | | | | | | | |
|---|----------------------|----------------------|----------------------|----------------------|----------------------|----------------------|-----------------------|
| Rate | 4 th Year | 5 th Year | 6 th Year | 7 th Year | 8 th Year | 9 th Year | 10 th Year |
| Intubating Video Laryngoscope with Macintosh Blade | | | | | | | |
| Labour | 19,243.21 | 21,167.53 | 23,284.28 | 25,612.71 | 28,173.98 | 30,991.38 | 34,090.52 |
| Comprehensive | 38,486.42 | 42,335.06 | 46,568.56 | 51,225.42 | 56,347.96 | 61,982.76 | 68,181.00 |

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Intubating Video Laryngoscope with Macintosh Blade

1. Should have Macintosh blades (adult large and extra large 1No each and Paediatric 1 No) with closed European metal finish with integrated camera chip and LED light illumination for obtaining more than 500000 Lux of brightness.
2. Should have Video blades of different sizes and handles constructed of durable, stainless steel and aluminium materials, and drop tested to 48 inches.

3. Should have Screen 7 inch or more in size for display with feature control buttons on the screen with HDMI output for connecting to a big screen. This screen should have option to connect to flexible intubating video endoscope cable as well.
4. Should have High definition and high resolution images features
5. Should have still picture and video recording capability via removable SD card for documentation required
6. Should have Integrated rechargeable battery with a backup of 2 hours.
7. A Soft bag should be supplied to place the monitor and system can also be operated without taking monitor out from the bag
8. IV stand for positioning of the monitor with tray for laryngoscope should be provided
9. Accessories like protection cap, tray for cleaning and sterilization of blades (at least two blades at a time) should be provided
10. All components should be sterilizable by standard methods.
11. Should provide Magill forceps for foreign body removal and for assisting nasal intubation
12. One imported trolley / box to hang scope as well monitor manufactured by same company should be provided.
13. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.