



Running Contract Details	
Equipment Name	Inverted microscope Model B
Running Contract Valid Till	18-07-2020
Tender Ref No	KMSCL/EP/T271/1109/2017(R)
Tendered Quantity	10
Supplier Name	M/s Bions Medical Systems Pvt Ltd
GST No	32AACCB4877B1Z1
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	5 Years

Supplier`s Details		
Address	Contact Details	
28/3085 D âœRohiniâœ Tagore Nagar Ponneth Temple Road Kadavanthra Kochi-682020.	Contact Person	Jitto John
	Phone	
	Mobile No	9567860501
	Email	bions@vsnl.net

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Inverted microscope Model B <i>Model & Make : DMIL LED Fluorescence / Leica Microsystems GmbH, Germany</i>	1547999.52 Incl.GST :18%	108359.97	1656359.49
2	Compatible adapter for attaching a Nikon DSLR camera <i>Model & Make : Leica</i>	103711.38 Incl.GST :18%	7259.8	110971.18
3	Camera with adapter <i>Model & Make : Leica</i>	392999 Incl.GST :18%	27509.93	420508.93
4	Additional filter I <i>Model & Make : Leica</i>	54200.94 Incl.GST :18%	3794.07	57995.01
5	Additional filter II <i>Model & Make : Leica</i>	66547.28 Incl.GST :18%	4658.31	71205.59
6	Additional filter III <i>Model & Make : Leica</i>	84944.66 Incl.GST :18%	5946.13	90890.79
		2250402.78	157528.19	2407930.97

Annual / Comprehensive Maintenance Charges (Exl.Tax)					
Rate	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Inverted microscope Model B					
Labour	29,500.00	30,975.00	32,524.34	34,150.38	35,857.84
Comprehensive	1,82,664.00	1,91,797.20	2,01,387.06	2,11,456.00	2,22,028.80

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Inverted microscope Model B

1. The equipment should be an inverted microscope with provision for BF, Phase contrast and fluorescence.
2. Should have wide field plan corrected 10X eyepiece of FOV 20mm OR direct monitor visualization.
3. Should have 4x, 10x, 20x and 40x long working distance full plan corrected objectives. Phase contrast and fluorescence should be available at 10x, 20x and 40x magnifications.
4. Condenser with a working distance of minimum 60mm.
5. Should have necessary stage attachments for holding 76mm microscopic slides, 25cm² flasks and 12 well, 24 well and 96 well flat bottom plates.
6. Should have filters for green and blue filters covering FITC, Acridine orange, Alexa 488, TRITC and option for adding additional filter.
7. Both bright field and fluorescence systems should have led illumination.
8. Should have provision for attaching a camera. The rate for compatible adapter for attaching a Nikon DSLR camera should be quoted separately. Also the rate for suitable camera with its adapter shall be quoted separately. OR should have an inbuilt camera. The camera model and make offered shall be provided in the offer form.
9. Should provide dust cover for stand
10. Should undertake calibration of the equipment as per the standards available.
11. Should provide four preventive maintenance visits per year and all spares including lamps and filters shall be replaced free of cost during warranty and CMC period.

12. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid