



Running Contract Details	
Equipment Name	Kerato Analyser
Running Contract Valid Till	09-04-2021
Tender Ref No	KMSCL/EP/T315/1306/2019(R)
Tendered Quantity	11
Supplier Name	M/s BioMedix Optotechnik & Devices Pvt Ltd
GST No	29AAACB7372B1ZU
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
#258 Amarjyothi Layout Domlur Banglore-560 071	Contact Person	Mr. William Dâ€™Mello , Muraleedhar,Arunkumar
	Phone	080-4933 2900
	Mobile No	9845248234,9611529900
	Email	info@biomedixdevices.com,sales@biomedix devices.com murali.sl@biomedixdevices.com salesupport@biomedixdevices.com info@biomedixdevices.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Kerato Analyser</b> <i>Model &amp; Make : Cell Check D / Konan Medical Inc</i>	2592800 Incl.GST :12%	191219	2784019
		<b>2592800</b>	<b>191219</b>	<b>2784019</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Kerato Analyser</b>							
Labour	35,000.00	40,000.00	45,000.00	50,000.00	55,000.00	60,000.00	65,000.00
Comprehensive	80,000.00	90,000.00	1,00,000.00	1,10,000.00	1,20,000.00	1,30,000.00	1,40,000.00

## **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

### **Equipment :Kerato Analyser**

1. Viewing field- 0.48 X 0.6mm(wide field)
2. Illumination – LED
3. Chamber compatibility- To all available systems
4. Camera- Dual CMOS(high resolution & High speed)
5. Display- LCD: Temperature & pachymetry
6. Vial dimension – Up to 35 mm
7. Range- XY & Z -16mmX20 mm
8. Thermometer- 0Ěšc to 40Ěšc
9. The system should be supplied with latest updated software.